

# Calibrating Unit New Standard Solution Report

## Equipment

Alcotest 7110 MKIII-C  
Location: SOUTH BRUNSWICK POLICE  
Serial No.: ARTL-0014  
Calibration File No.: 00406  
Certification File No.: 00407  
Linearity File No.: 00408  
Solution File No.: 00418  
Sequential File No.: 00418  
Calib. Date: 09/30/2014  
Cert. Date: 09/30/2014  
Lin. Date: 09/30/2014  
Soln. Date: 02/16/2015  
File Date: 02/16/2015  
Calib. No.: 00029  
Cert. No.: 00021  
Lin. No.: 00021  
Soln. No.: 00151

Calibrating Unit: WET  
Control Solution %: 0.100%  
Solution Control Lot: 14H132  
Model No.: CU-34  
Serial No.: DDUH S3-0079  
Expires: 08/22/2016  
Bottle No.: 0035

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:45S		
Control 1 EC	0.099%	13:46S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	13:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:47S		
Control 2 EC	0.098%	13:48S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	13:48S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:49S		
Control 3 EC	0.096%	13:49S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	13:49S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:50S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature:  #421

Badge No.: 421

Date: 02/16/2015