Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C		Serial No.: ARNK-0074
Location:	SOUTH BRUNSWICK POLICE		
Calibration File No.:	00831	Calib. Date: 02/17/2016	Calib. No.: 00032
Certification File No.:	00832	Cert. Date: 02/17/2016	Cert. No.: 00022
Linearity File No.:	00833	Lin. Date: 02/17/2016	Lin. No.: 00022
Solution File No.:	00850	Soln. Date: 05/12/2016	Soln. No.: 00179
Sequential File No.:	00850	File Date: 05/12/2016	

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDUF S3-0061
Control Solution %:	0.100%		Expires: 08/04/2017
Solution Control Lot:	15H140		Bottle No.: 0088

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	18:11D		
Control 1 EC	0.101%	18:11D	34.1°C	*** TEST PASSED ***
Control 1 IR	0.101%	18:11D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:12D		
Control 2 EC	0.101%	18:13D	34.1°C	*** TEST PASSED ***
Control 2 IR	0.100%	18:13D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:14D		
Control 3 EC	0.100%	18:14D	34.1°C	*** TEST PASSED ***
Control 3 IR	0.100%	18:14D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:15D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE First Name: JESSE MI: R

Signature: PH #421

Badge No.: 421

Date: 05/12/2016