Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARTL-0014
Location:	SOUTH BRUNSWICK POLICE				
Calibration File No.:	00257	Calib. Date:	06/21/2010	Calib. No.:	00016
Certification File No.:	00258	Cert. Date:	06/21/2010	Cert. No.:	00010
Linearity File No.:	00259	Lin. Date:	06/21/2010	Lin. No.:	00010
Solution File No.:	00264	Soln. Date:	07/20/2010	Soln. No.:	00076
Sequential File No.:	00264	File Date:	07/20/2010		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUH S3-0079
Control Solution %:	0.100%			Expires:	09/18/2011
Solution Control Lot:	09I069			Bottle No.:	0376

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:17D		
Control 1 EC	0.098%	10:18D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	10:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:18D		
Control 2 EC	0.097%	10:19D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:20D		
Control 3 EC	0.098%	10:20D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	10:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:21D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ROGERS First Name: MICHAEL MI: J.

Signature: PFC Michael J. Rogers

Badge No.: 258

Date: 07/20/2010