

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0014
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 00221 Calib. Date: 06/24/2009 Calib. No.: 00010
Certification File No.: 00222 Cert. Date: 06/24/2009 Cert. No.: 00008
Linearity File No.: 00223 Lin. Date: 06/24/2009 Lin. No.: 00008
Solution File No.: 00225 Soln. Date: 07/22/2009 Soln. No.: 00061
Sequential File No.: 00225 File Date: 07/22/2009

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0060
Control Solution %: 0.100% Expires: 12/01/2010
Solution Control Lot: 08L061 Bottle No.: 1074

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:39D		
Control 1 EC	0.102%	12:40D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	12:40D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:40D		
Control 2 EC	0.101%	12:41D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:42D		
Control 3 EC	0.101%	12:42D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	12:42D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:43D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: ROGERS

First Name: MICHAEL

MI: J.

Signature: _____



Badge No.: 258

Date: 07/22/2009