

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0014
Location:	SOUTH BRUNSWICK POLICE		
Calibration File No.:	00287	Calib. Date:	03/22/2011
Certification File No.:	00288	Cert. Date:	03/22/2011
Linearity File No.:	00289	Lin. Date:	03/22/2011
Solution File No.:	00304	Soln. Date:	08/02/2011
Sequential File No.:	00304	File Date:	08/02/2011
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUH S3-0079
Solution Control Lot:	10L084	Expires:	12/16/2012
		Bottle No.:	0143

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:41D		
Control 1 EC	0.102%	09:41D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	09:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:42D		
Control 2 EC	0.102%	09:42D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.103%	09:42D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:43D		
Control 3 EC	0.102%	09:43D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.103%	09:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:44D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: _____



#351

Badge No.: 351

Date: 08/02/2011