

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00394	Calib. Date: 05/01/2014
Certification File No.:	00395	Calib. No.: 00028
Linearity File No.:	00396	Cert. Date: 05/01/2014
Solution File No.:	00402	Cert. No.: 00020
Sequential File No.:	00402	Lin. Date: 05/01/2014
		Lin. No.: 00020
		Soln. Date: 08/14/2014
		Soln. No.: 00141
		File Date: 08/14/2014
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUH S3-0079
Solution Control Lot:	13K125	Expires: 11/12/2015
		Bottle No.: 1336

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:12D		
Control 1 EC	0.099%	13:13D	34.0°C	AMBIENT AIR CHECK ERR
Control 1 IR	-.---%	13:13D	34.0°C	AMBIENT AIR CHECK ERR
Ambient Air Blank	-.---%	13:14D		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____

Badge No.: 421

Date: 08/14/2014

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	Calib. No.:	00028
Calibration File No.:	00394	Cert. Date:	05/01/2014
Certification File No.:	00395	Lin. Date:	05/01/2014
Linearity File No.:	00396	Soln. Date:	08/14/2014
Solution File No.:	00403	File Date:	08/14/2014
Sequential File No.:	00403		
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUH S3-0079
Solution Control Lot:	13K125	Expires:	11/12/2015
		Bottle No.:	1336

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:44D		
Control 1 EC	0.099%	14:45D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	14:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:45D		
Control 2 EC	0.098%	14:46D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	14:46D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:46D		
Control 3 EC	0.098%	14:47D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	14:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:48D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____



Badge No.: 421

Date: 08/14/2014