Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C		Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE		
Calibration File No.:	00327	Calib. Date: 07/02/2012	Calib. No.: 00023
Certification File No.:	00328	Cert. Date: 07/02/2012	Cert. No.: 00015
Linearity File No.:	00329	Lin. Date: 07/02/2012	Lin. No.: 00015
Solution File No.:	00333	Soln. Date: 08/29/2012	Soln. No.: 00111
Sequential File No.:	00333	File Date: 08/29/2012	

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDUH S3-0079
Control Solution %:	0.100%		Expires: 09/15/2013
Solution Control Lot:	11I095		Bottle No.: 0446

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	18:33D		
Control 1 EC	0.099%	18:34D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	18:34D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:35D		
Control 2 EC	0.099%	18:35D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	18:35D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:36D		
Control 3 EC	0.098%	18:36D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	18:36D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:37D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN First Name: GARY MI: W

Signature: #351 Badge No.: 351

Date: 08/29/2012

Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C		Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE		
Calibration File No.:	00327	Calib. Date: 07/02/2012	Calib. No.: 00023
Certification File No.:	00328	Cert. Date: 07/02/2012	Cert. No.: 00015
Linearity File No.:	00329	Lin. Date: 07/02/2012	Lin. No.: 00015
Solution File No.:	00332	Soln. Date: 08/28/2012	Soln. No.: 00110
Sequential File No.:	00332	File Date: 08/28/2012	

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDUH S3-0079
Control Solution %:	0.100%		Expires: 09/15/2013
Solution Control Lot:	11I095		Bottle No.: 0442

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	18:58D		
Control 1 EC	0.099%	18:59D	33.9°C	AMBIENT AIR CHECK ERR
Control 1 IR	-.---%	18:59D	33.9°C	AMBIENT AIR CHECK ERR
Ambient Air Blank	-.---%	18:59D		
Control 2 EC	-.---%	--:---	N/A	
Control 2 IR	-.---%	--:---	N/A	
Ambient Air Blank	-.---%	--:---		
Control 3 EC	-.---%	--:---	N/A	
Control 3 IR	-.---%	--:---	N/A	
Ambient Air Blank	-.---%	--:---		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN First Name: GARY MI: W

Signature: H351 Badge No.: 351

Date: 08/28/2012