

Calibrating Unit New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: SOUTH BRUNSWICK POLICE
Serial No.: ARNK-0074
Calibration File No.: 00673 Calib. Date: 04/07/2014 Calib. No.: 00028
Certification File No.: 00674 Cert. Date: 04/07/2014 Cert. No.: 00018
Linearity File No.: 00675 Lin. Date: 04/07/2014 Lin. No.: 00018
Solution File No.: 00702 Soln. Date: 09/15/2014 Soln. No.: 00154
Sequential File No.: 00702 File Date: 09/15/2014

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0061
Control Solution %: 0.100% Expires: 11/12/2015
Solution Control Lot: 13K125 Bottle No.: 1332

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:00D		
Control 1 EC	0.099%	08:01D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	08:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:01D		
Control 2 EC	0.098%	08:02D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	08:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:02D		
Control 3 EC	0.098%	08:03D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	08:03D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:04D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____



Badge No.: 421

Date: 09/15/2014