Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARTL-0014
Location:	SOUTH BRUNSWICK POLICE				
Calibration File No.:	00305	Calib. Date:	08/26/2011	Calib. No.:	00021
Certification File No.:	00306	Cert. Date:	08/26/2011	Cert. No.:	00013
Linearity File No.:	00307	Lin. Date:	08/26/2011	Lin. No.:	00013
Solution File No.:	00309	Soln. Date:	09/23/2011	Soln. No.:	00094
Sequential File No.:	00309	File Date:	09/23/2011		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUH S3-0079
Control Solution %:	0.100%			Expires:	12/16/2012
Solution Control Lot:	10L084			Bottle No.:	0147

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:51D		
Control 1 EC	0.100%	10:51D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:51D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:52D		
Control 2 EC	0.099%	10:52D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	10:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:53D		
Control 3 EC	0.099%	10:53D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	10:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:54D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN First Name: GARY MI: W

Signature: #351

Badge No.: 351

Date: 09/23/2011