

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARNK-0074  
**Location:** SOUTH BRUNSWICK POLICE  
**Calibration File No.:** 01143 **Calib. Date:** 12/31/2020 **Calib. No.:** 00045  
**Certification File No.:** 01144 **Cert. Date:** 12/31/2020 **Cert. No.:** 00033  
**Linearity File No.:** 01145 **Lin. Date:** 12/31/2020 **Lin. No.:** 00033  
**Solution File No.:** 01158 **Soln. Date:** 03/16/2021 **Soln. No.:** 00251  
**Sequential File No.:** 01158 **File Date:** 03/16/2021

**Calibrating Unit:** WET **Model No.:** CU-34 **Serial No.:** DDUF S3-0061  
**Control Solution %:** 0.100% **Expires:** 02/19/2022  
**Solution Control Lot:** 20090 **Bottle No.:** 0010

| Function          | Result<br>%BAC | Time<br>HH:MM | Temperature<br>Simulator (°C) | Comment(s)<br>or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000%         | 10:45S        |                               |                           |
| Control 1 EC      | 0.099%         | 10:45S        | 34.0°C                        | *** TEST PASSED ***       |
| Control 1 IR      | 0.100%         | 10:45S        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 10:46S        |                               |                           |
| Control 2 EC      | 0.099%         | 10:47S        | 34.0°C                        | *** TEST PASSED ***       |
| Control 2 IR      | 0.099%         | 10:47S        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 10:47S        |                               |                           |
| Control 3 EC      | 0.100%         | 10:48S        | 34.0°C                        | *** TEST PASSED ***       |
| Control 3 IR      | 0.099%         | 10:48S        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 10:49S        |                               |                           |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: \_\_\_\_\_

*[Handwritten Signature]* #421

Badge No.: 421

Date: 03/16/2021