

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0074
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	01123	Calib. Date: 07/14/2020
Certification File No.:	01124	Calib. No.: 00044
Linearity File No.:	01125	Cert. Date: 07/14/2020
Solution File No.:	01135	Cert. No.: 00032
Sequential File No.:	01135	Lin. Date: 07/14/2020
		Lin. No.: 00032
		Soln. Date: 10/02/2020
		Soln. No.: 00244
		File Date: 10/02/2020
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUF S3-0060
Solution Control Lot:	19020	Expires: 01/08/2021
		Bottle No.: 0335

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:40D		
Control 1 EC	0.100%	07:41D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	07:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:42D		
Control 2 EC	0.099%	07:42D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	07:42D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:43D		
Control 3 EC	0.100%	07:43D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	07:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:44D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____

PH  #421

Badge No.: 421

Date: 10/02/2020