

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0074
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 01162 Calib. Date: 06/16/2021 Calib. No.: 00047
Certification File No.: 01163 Cert. Date: 06/16/2021 Cert. No.: 00034
Linearity File No.: 01164 Lin. Date: 06/16/2021 Lin. No.: 00034
Solution File No.: 01174 Soln. Date: 08/26/2021 Soln. No.: 00256
Sequential File No.: 01174 File Date: 08/26/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUH S3-0080
Control Solution %: 0.100% Expires: 01/13/2023
Solution Control Lot: 21010 Bottle No.: 1428

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:27D		
Control 1 EC	0.100%	10:28D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	10:28D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:28D		
Control 2 EC	0.100%	10:29D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	10:29D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:30D		
Control 3 EC	0.100%	10:30D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	10:30D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:31D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____

Jesse Blake #421

Badge No.: 421

Date: 08/26/2021