

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARNK-0074  
Location: SOUTH BRUNSWICK POLICE  
Calibration File No.: 01177 Calib. Date: 09/09/2021 Calib. No.: 00048  
Certification File No.: 01178 Cert. Date: 09/09/2021 Cert. No.: 00035  
Linearity File No.: 01179 Lin. Date: 09/09/2021 Lin. No.: 00035  
Solution File No.: 01183 Soln. Date: 09/30/2021 Soln. No.: 00258  
Sequential File No.: 01183 File Date: 09/30/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUH S3-0079  
Control Solution %: 0.100% Expires: 01/13/2023  
Solution Control Lot: 21010 Bottle No.: 1432

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:51D		
Control 1 EC	0.100%	08:52D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	08:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:52D		
Control 2 EC	0.100%	08:53D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.102%	08:53D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:54D		
Control 3 EC	0.100%	08:54D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	08:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:55D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**  
Last Name: BLAKE First Name: JESSE MI: R  
Signature:  #421 Badge No.: 421  
Date: 09/30/2021