Calibrating Unit
New Standard Solution Report

Equipment Alcotest 7110 MKIII-C — Serial No.: ARTL-0014

Location: SOUTH BRUNSWICK POLICE

Calibration File No.:	00606	Calib. Date:	03/31/2021	Calib. No.:	00047
Certification File No.:	00607	Cert. Date:	03/31/2021	Cert. No.:	00036
Linearity File No.:	00608	Lin. Date:	03/31/2021	Lin. No.:	00036
Solution File No.:	00620	Soln. Date:	06/23/2021	Soln. No.:	00252
Sequential File No.:	00620	File Date:	06/23/2021		

Calibrating Unit: WET — Model No.: CU-34 — Serial No.: DDUF S3-0060

Control Solution %: 0.100% — Expires: 01/13/2023

Solution Control Lot: 21010 — Bottle No.: 1270

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:09D		
Control 1 EC	0.101%	08:10D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	08:10D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:11D		
Control 2 EC	0.101%	08:11D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	08:11D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:12D		
Control 3 EC	0.101%	08:12D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	08:12D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:13D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE — First Name: JESSE — MI: R

Signature: Pfl. #421 — Badge No.: 421

Date: 06/23/2021