

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARNK-0074  
Location: SOUTH BRUNSWICK POLICE  
Calibration File No.: 01143 Calib. Date: 12/31/2020 Calib. No.: 00045  
Certification File No.: 01144 Cert. Date: 12/31/2020 Cert. No.: 00033  
Linearity File No.: 01145 Lin. Date: 12/31/2020 Lin. No.: 00033  
Solution File No.: 01150 Soln. Date: 01/28/2021 Soln. No.: 00249  
Sequential File No.: 01150 File Date: 01/28/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0061  
Control Solution %: 0.100% Expires: 02/19/2022  
Solution Control Lot: 20090 Bottle No.: 0396

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:50S		
Control 1 EC	0.098%	08:51S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	08:51S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:52S		
Control 2 EC	0.098%	08:52S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	08:52S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:53S		
Control 3 EC	0.099%	08:54S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	08:54S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:55S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: \_\_\_\_\_

*Jesse Blake #421*

Badge No.: 421

Date: 01/28/2021