

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0074
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 01210 Calib. Date: 02/10/2022 Calib. No.: 00049
Certification File No.: 01211 Cert. Date: 02/10/2022 Cert. No.: 00036
Linearity File No.: 01212 Lin. Date: 02/10/2022 Lin. No.: 00036
Solution File No.: 01222 Soln. Date: 03/08/2022 Soln. No.: 00264
Sequential File No.: 01222 File Date: 03/08/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUH S3-0080
Control Solution %: 0.100% Expires: 01/13/2023
Solution Control Lot: 21010 Bottle No.: 1283

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:48S		
Control 1 EC	0.099%	09:49S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	09:49S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:50S		
Control 2 EC	0.099%	09:50S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:50S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:51S		
Control 3 EC	0.099%	09:52S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:52S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:53S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____



Badge No.: 421

Date: 03/08/2022