

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARNK-0074  
**Location:** SOUTH BRUNSWICK POLICE  
**Calibration File No.:** 01210 **Calib. Date:** 02/10/2022 **Calib. No.:** 00049  
**Certification File No.:** 01211 **Cert. Date:** 02/10/2022 **Cert. No.:** 00036  
**Linearity File No.:** 01212 **Lin. Date:** 02/10/2022 **Lin. No.:** 00036  
**Solution File No.:** 01229 **Soln. Date:** 04/05/2022 **Soln. No.:** 00265  
**Sequential File No.:** 01229 **File Date:** 04/05/2022

**Calibrating Unit:** WET **Model No.:** CU-34 **Serial No.:** DDUH S3-0080  
**Control Solution %:** 0.100% **Expires:** 09/01/2023  
**Solution Control Lot:** 21360 **Bottle No.:** 1201

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:51D		
Control 1 EC	0.098%	08:51D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	08:51D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:52D		
Control 2 EC	0.098%	08:53D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	08:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:53D		
Control 3 EC	0.098%	08:54D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	08:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:55D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: \_\_\_\_\_

*Pfc. Jesse Blake #421*

Badge No.: 421

Date: 04/05/2022