

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARNK-0074  
**Location:** SOUTH BRUNSWICK POLICE  
**Calibration File No.:** 01053 **Calib. Date:** 07/10/2019 **Calib. No.:** 00042  
**Certification File No.:** 01054 **Cert. Date:** 07/10/2019 **Cert. No.:** 00030  
**Linearity File No.:** 01055 **Lin. Date:** 07/10/2019 **Lin. No.:** 00030  
**Solution File No.:** 01056 **Soln. Date:** 07/10/2019 **Soln. No.:** 00227  
**Sequential File No.:** 01056 **File Date:** 07/10/2019

**Calibrating Unit:** WET **Model No.:** CU-34 **Serial No.:** DDUF S3-0060  
**Control Solution %:** 0.100% **Expires:** 05/10/2020  
**Solution Control Lot:** 18150 **Bottle No.:** 1231

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:18D		
Control 1 EC	0.100%	10:18D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:18D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:19D		
Control 2 EC	0.100%	10:20D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:20D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:20D		
Control 3 EC	0.100%	10:21D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	10:21D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:22D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDEE P2-034 SA

**Changed By:**

Last Name: MIMIKOS

First Name: NICHOLAS

MI: E.

Signature: TPR. N.A. #7413

Badge No.: 7413

Date: 07/10/2019