

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0074
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 01177 Calib. Date: 09/09/2021 Calib. No.: 00048
Certification File No.: 01178 Cert. Date: 09/09/2021 Cert. No.: 00035
Linearity File No.: 01179 Lin. Date: 09/09/2021 Lin. No.: 00035
Solution File No.: 01204 Soln. Date: 01/17/2022 Soln. No.: 00262
Sequential File No.: 01204 File Date: 01/17/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUH S3-0079
Control Solution %: 0.100% Expires: 01/13/2023
Solution Control Lot: 21010 Bottle No.: 1271

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:37S		
Control 1 EC	0.100%	08:38S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	08:38S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:39S		
Control 2 EC	0.100%	08:40S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	08:40S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:40S		
Control 3 EC	0.101%	08:41S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	08:41S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:42S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MARRERO

First Name: JOSEPH

MI: A

Signature: _____



406

Badge No.: 406

Date: 01/17/2022