Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARNK-0074
Location:	SOUTH BRUNSWICK POLICE				
Calibration File No.:	01123	Calib. Date:	07/14/2020	Calib. No.:	00044
Certification File No.:	01124	Cert. Date:	07/14/2020	Cert. No.:	00032
Linearity File No.:	01125	Lin. Date:	07/14/2020	Lin. No.:	00032
Solution File No.:	01140	Soln. Date:	12/18/2020	Soln. No.:	00247
Sequential File No.:	01140	File Date:	12/18/2020		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUF S3-0060
Control Solution %:	0.100%			Expires:	02/19/2022
Solution Control Lot:	20090			Bottle No.:	0365

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:21S		
Control 1 EC	0.100%	07:21S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	07:21S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:22S		
Control 2 EC	0.099%	07:23S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	07:23S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:24S		
Control 3 EC	0.099%	07:24S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	07:24S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:25S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE First Name: JESSE MI: R

Signature: #421 Badge No.: 421

Date: 12/18/2020