

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0074
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	01143	Calib. Date: 12/31/2020
Certification File No.:	01144	Calib. No.: 00045
Linearity File No.:	01145	Cert. Date: 12/31/2020
Solution File No.:	01154	Cert. No.: 00033
Sequential File No.:	01154	Lin. Date: 12/31/2020
		Lin. No.: 00033
		Soln. Date: 02/25/2021
		Soln. No.: 00250
		File Date: 02/25/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUF S3-0061
Solution Control Lot:	20090	Expires: 02/19/2022
		Bottle No.: 0263

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:05S		
Control 1 EC	0.098%	08:05S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:05S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:06S		
Control 2 EC	0.099%	08:07S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:07S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:07S		
Control 3 EC	0.099%	08:08S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	08:08S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:09S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____

Badge No.: 421

Date: 02/25/2021