

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0074
Location:	SOUTH BRUNSWICK POLICE		
Calibration File No.:	01162	Calib. Date:	06/16/2021
Certification File No.:	01163	Cert. Date:	06/16/2021
Linearity File No.:	01164	Lin. Date:	06/16/2021
Solution File No.:	01171	Soln. Date:	08/04/2021
Sequential File No.:	01171	File Date:	08/04/2021
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUH S3-0080
Solution Control Lot:	21010	Expires:	01/13/2023
		Bottle No.:	1282

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:38D		
Control 1 EC	0.101%	08:38D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:39D		
Control 2 EC	0.100%	08:40D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	08:40D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:40D		
Control 3 EC	0.100%	08:41D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	08:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:41D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____



Badge No.: 421

Date: 08/04/2021