

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARNK-0074  
Location: SOUTH BRUNSWICK POLICE  
Calibration File No.: 01123 Calib. Date: 07/14/2020 Calib. No.: 00044  
Certification File No.: 01124 Cert. Date: 07/14/2020 Cert. No.: 00032  
Linearity File No.: 01125 Lin. Date: 07/14/2020 Lin. No.: 00032  
Solution File No.: 01131 Soln. Date: 09/04/2020 Soln. No.: 00243  
Sequential File No.: 01131 File Date: 09/04/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0060  
Control Solution %: 0.100% Expires: 01/08/2021  
Solution Control Lot: 19020 Bottle No.: 0279

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:28D		
Control 1 EC	0.100%	10:29D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:29D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:29D		
Control 2 EC	0.099%	10:30D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:30D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:31D		
Control 3 EC	0.100%	10:31D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:31D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:32D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: \_\_\_\_\_



Badge No.: 421

Date: 09/04/2020