

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0074
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 01162 **Calib. Date:** 06/16/2021 **Calib. No.:** 00047
Certification File No.: 01163 **Cert. Date:** 06/16/2021 **Cert. No.:** 00034
Linearity File No.: 01164 **Lin. Date:** 06/16/2021 **Lin. No.:** 00034
Solution File No.: 01169 **Soln. Date:** 07/14/2021 **Soln. No.:** 00254
Sequential File No.: 01169 **File Date:** 07/14/2021

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDUH S3-0080
Control Solution %: 0.100% **Expires:** 01/13/2023
Solution Control Lot: 21010 **Bottle No.:** 1429

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:04D		
Control 1 EC	0.100%	08:05D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:05D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:05D		
Control 2 EC	0.100%	08:06D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	08:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:06D		
Control 3 EC	0.100%	08:07D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	08:07D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:08D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____

PC  #421

Badge No.: 421

Date: 07/14/2021