

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0014
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 00570 Calib. Date: 01/25/2020 Calib. No.: 00042
Certification File No.: 00571 Cert. Date: 01/25/2020 Cert. No.: 00033
Linearity File No.: 00572 Lin. Date: 01/25/2020 Lin. No.: 00033
Solution File No.: 00578 Soln. Date: 05/05/2020 Soln. No.: 00235
Sequential File No.: 00578 File Date: 05/05/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUH S3-0079
Control Solution %: 0.100% Expires: 02/19/2022
Solution Control Lot: 20090 Bottle No.: 0171

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	17:12D		
Control 1 EC	0.103%	17:13D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	17:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:14D		
Control 2 EC	0.100%	17:15D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	17:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:16D		
Control 3 EC	0.101%	17:17D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	17:17D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:18D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: 

Badge No.: 351

Date: 05/05/2020