Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARTL-0014
Location:	SOUTH BRUNSWICK POLICE				
Calibration File No.:	00546	Calib. Date:	01/14/2019	Calib. No.:	00040
Certification File No.:	00547	Cert. Date:	01/14/2019	Cert. No.:	00031
Linearity File No.:	00548	Lin. Date:	01/14/2019	Lin. No.:	00031
Solution File No.:	00559	Soln. Date:	05/30/2019	Soln. No.:	00223
Sequential File No.:	00559	File Date:	05/30/2019		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUH S3-0079
Control Solution %:	0.100%			Expires:	05/10/2020
Solution Control Lot:	18150			Bottle No.:	1296

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:46D		
Control 1 EC	0.105%	10:47D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	10:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:47D		
Control 2 EC	0.102%	10:48D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	10:48D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:49D		
Control 3 EC	0.102%	10:49D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	10:49D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:50D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE First Name: JESSE MI: R

Signature: PH. #421 Badge No.: 421

Date: 05/30/2019

Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARTL-0014
Location:	SOUTH BRUNSWICK POLICE				
Calibration File No.:	00546	Calib. Date:	01/14/2019	Calib. No.:	00040
Certification File No.:	00547	Cert. Date:	01/14/2019	Cert. No.:	00031
Linearity File No.:	00548	Lin. Date:	01/14/2019	Lin. No.:	00031
Solution File No.:	00558	Soln. Date:	05/30/2019	Soln. No.:	00222
Sequential File No.:	00558	File Date:	05/30/2019		
Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUH S3-0079
Control Solution %:	0.100%			Expires:	05/10/2020
Solution Control Lot:	18150			Bottle No.:	1296

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:38D		
Control 1 EC	-.---%	07:38D	22.8°C	SIMULATOR TEMP. ERROR
Control 1 IR	-.---%	07:38D	22.8°C	SIMULATOR TEMP. ERROR
Ambient Air Blank	-.---%	--:---		
Control 2 EC	-.---%	--:---	N/A	
Control 2 IR	-.---%	--:---	N/A	
Ambient Air Blank	-.---%	--:---		
Control 3 EC	-.---%	--:---	N/A	
Control 3 IR	-.---%	--:---	N/A	
Ambient Air Blank	-.---%	--:---		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE First Name: JESSE MI: R

Signature: _____ Badge No.: 421

Date: 05/30/2019