Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARTL-0014
Location:	SOUTH BRUNSWICK POLICE				
Calibration File No.:	00584	Calib. Date:	07/14/2020	Calib. No.:	00043
Certification File No.:	00585	Cert. Date:	07/14/2020	Cert. No.:	00034
Linearity File No.:	00586	Lin. Date:	07/14/2020	Lin. No.:	00034
Solution File No.:	00591	Soln. Date:	09/10/2020	Soln. No.:	00242
Sequential File No.:	00591	File Date:	09/10/2020		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUH S3-0080
Control Solution %:	0.100%			Expires:	02/19/2022
Solution Control Lot:	20090			Bottle No.:	0229

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:01D		
Control 1 EC	0.101%	14:02D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	14:02D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:02D		
Control 2 EC	0.099%	14:03D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.096%	14:03D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:04D		
Control 3 EC	0.098%	14:04D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.095%	14:04D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:05D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE First Name: JESSE MI: R

Signature: PTI #421 Badge No.: 421

Date: 09/10/2020