

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARNK-0074  
Location: SOUTH BRUNSWICK POLICE  
Calibration File No.: 01210 Calib. Date: 02/10/2022 Calib. No.: 00049  
Certification File No.: 01211 Cert. Date: 02/10/2022 Cert. No.: 00036  
Linearity File No.: 01212 Lin. Date: 02/10/2022 Lin. No.: 00036  
Solution File No.: 01244 Soln. Date: 05/31/2022 Soln. No.: 00267  
Sequential File No.: 01244 File Date: 05/31/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUH S3-0080  
Control Solution %: 0.100% Expires: 09/01/2023  
Solution Control Lot: 21360 Bottle No.: 0966

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:33D		
Control 1 EC	0.098%	08:33D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	08:33D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:34D		
Control 2 EC	0.097%	08:35D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	08:35D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:35D		
Control 3 EC	0.098%	08:36D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	08:36D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:37D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: GASSMAN

First Name: JASON

MI:

Signature: 

Badge No.: 408

Date: 05/31/2022