Calibrating Unit
New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0014
Location: SOUTH BRUNSWICK POLICE

Calibration File No.:	00658	Calib. Date:	06/28/2022	Calib. No.:	00050
Certification File No.:	00659	Cert. Date:	06/28/2022	Cert. No.:	00039
Linearity File No.:	00660	Lin. Date:	06/28/2022	Lin. No.:	00039
Solution File No.:	00669	Soln. Date:	08/23/2022	Soln. No.:	00269
Sequential File No.:	00669	File Date:	08/23/2022		

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0061
Control Solution %: 0.100% Expires: 09/01/2023
Solution Control Lot: 21360 Bottle No.: 1106

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:46D		
Control 1 EC	0.100%	07:47D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	07:47D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:48D		
Control 2 EC	0.100%	07:48D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	07:48D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:49D		
Control 3 EC	0.101%	07:49D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	07:49D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:50D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:
Last Name: GASSMAN First Name: JASON MI:
Signature: PFC J M 408 Badge No.: 408
Date: 08/23/2022