

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0014
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 00658 **Calib. Date:** 06/28/2022 **Calib. No.:** 00050
Certification File No.: 00659 **Cert. Date:** 06/28/2022 **Cert. No.:** 00039
Linearity File No.: 00660 **Lin. Date:** 06/28/2022 **Lin. No.:** 00039
Solution File No.: 00678 **Soln. Date:** 10/20/2022 **Soln. No.:** 00271
Sequential File No.: 00678 **File Date:** 10/20/2022

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDUF S3-0061
Control Solution %: 0.100% **Expires:** 09/01/2023
Solution Control Lot: 21360 **Bottle No.:** 0306

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:21D		
Control 1 EC	0.100%	11:22D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:22D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:22D		
Control 2 EC	0.100%	11:23D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:23D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:24D		
Control 3 EC	0.100%	11:24D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:24D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:25D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GASSMAN

First Name: JASON

MI:

Signature: _____

PFC J M 408

Badge No.: 408

Date: 10/20/2022