Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARNK-0074
Location:	SOUTH BRUNSWICK POLICE				
Calibration File No.:	01316	Calib. Date:	06/28/2023	Calib. No.:	00053
Certification File No.:	01317	Cert. Date:	06/28/2023	Cert. No.:	00040
Linearity File No.:	01318	Lin. Date:	06/28/2023	Lin. No.:	00040
Solution File No.:	01326	Soln. Date:	08/16/2023	Soln. No.:	00288
Sequential File No.:	01326	File Date:	08/16/2023		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUH S3-0079
Control Solution %:	0.100%			Expires:	01/11/2025
Solution Control Lot:	23010			Bottle No.:	0940

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:06D		
Control 1 EC	0.103%	07:07D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.102%	07:07D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:08D		
Control 2 EC	0.102%	07:08D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.102%	07:08D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:09D		
Control 3 EC	0.101%	07:09D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.102%	07:09D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:10D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GASSMAN First Name: JASO MI:

Signature: PFC JM 408

Badge No.: 408

Date: 08/16/2023