

ALCOTEST 9510 PARAMETER REPORT

Equipment

Serial No.: ARMK-0097
Firmware: 8326739 1.5
WinCE application: 8326738 2.9
Configuration: 8326737 3.10

Date: 01/15/2026
Time: 08:45:06

Parameter

min. blow time	5.0	s
min. breath volume for females of age 60+	1.2	L
min. breath volume for all other	1.5	L
min. blow flow	4.5	L/min
plateau detection limit	4	%
plateau detection start conc.	70	microgram/L
neg. flow detection (part. vacuum)	10.0	hPa
neg. flow detection sensitivity	10	
cal. gas abort volume	0.4	L
result-to-zero limit	0.0050	%BAC
ambient air check limit	0.0049	%BAC
interference det. d-criterion limit abs.	38	microgram/L
interference det. d-criterion limit rel.	10.0	%
interference det. t-criterion limit abs.	8	microgram/L
interference det. t-criterion limit rel.	2.1	%
IR CO2 offset	10	microgram/L
IR H2O offset	4	microgram/L
EC H2O offset	0	microgram/L
Value-based EC aging comp. on/off (1/0)	0	
Time-based EC aging comp. on/off (1/0)	1	
Time-based EC aging comp. per month	0.2	%
Time-based EC aging comp. maximum	3.0	%
EC fatigue comp. max. sum	15000	
EC fatigue comp. factor	50	
EC fatigue comp. minutes	180	
mouth alc. mark limit	500	
mouth alc. lower limit	30	
mouth alc. slope	6	
mouth alc. zero limit	50	
mouth alc. max. neg. sum	6	
mouth alc. max. 2nd derivative	35	

ALCOTEST 9510 CERTIFICATION REPORT - WET ADJUST (PART I)
South Brunswick

Equipment

Inst. Model No.: ALCOTEST 9510 Serial No.: ARMK-0097
Firmware: 8326739 1.5 Config.: 8326737 3.10 WinCE: 8326738 2.9

Wet Adjust Record

Wet Adjust File No.: 108 Wet Adjust Date: 01/15/2026 Wet Adjust No.: 7
Wet Adjust Time: 09:38:57
Concentration: 0.100 %
Adjusting Unit: X-Cal 2000 Adj. Unit Ser. No.: ARND-0002 Adj. Unit Exp.: 08/20/2026
Solution Lot No.: 24210 Soln. Bottle No.: 618 Adjust Soln. Exp.: 06/11/2026
Preadjust Simulator Temp.: 34.01 degree C
Postadjust Simulator Temp.: 34.01 degree C

Result

Procedure completed successfully.

Coordinator

Last Name: Fillimon - First Name: James MI: G Badge No.: 8306

On this date, I certified the above instrument in accordance with the Alcotest 9510 operator training and procedures established by the NJSP Office of Forensic Sciences.

TP. James Fillimon # 8306

Signed:

Date: 01/15/2026

ID: 2673

ALCOTEST 9510 CERTIFICATION REPORT - DRY ADJUST (PART II)
South Brunswick

Equipment

Inst. Model No.: ALCOTEST 9510 Serial No.: ARMK-0097
Firmware: 8326739 1.5 Config.: 8326737 3.10 WinCE: 8326738 2.9

Dry Adjust Record

Dry Adjust File No.: 109 Dry Adjust Date: 01/15/2026 Dry Adjust No.: 7
Dry Adjust Time: 09:57:41

Concentration: 0.100 %
Dry Gas Lot No.: 302-403034216 Adjust Gas Exp.: 04/30/2027
Barom. Model No.: Mensor CPG2300 Barom. Serial No.: 41001272 Barom. Cert. Exp.: 08/25/2026
Pre-adjust Amb. Pressure: 996 hPa Post-adjust Amb. Pressure: 995 hPa

Result

Procedure completed successfully.

Coordinator

Last Name: Fillimon - First Name: James MI: G Badge No.: 8306

On this date, I certified the above instrument in accordance with the Alcotest 9510 operator training and procedures established by the NJSP Office of Forensic Sciences.

Tr. James Fillimon # 8306

Signed:

Date: 01/15/2026

ID: 2673

ALCOTEST 9510 CERTIFICATION REPORT - LINEARITY (PART III)
South Brunswick

Equipment

Inst. Model No.: ALCOTEST 9510 Serial No.: ARMK-0097
 Firmware: 8326739 1.5 Config.: 8326737 3.10 WinCE: 8326738 2.9

Linearity Record

Linearity File No.: 110 Lin. Date: 01/15/2026 Lin. No.: 7

0.040% Dry Gas Lot No.: 302-402730318 Adjust. Gas Exp.: 04/27/2026
 0.080% Dry Gas Lot No.: 302-402732434 Adjust. Gas Exp.: 04/28/2026
 0.160% Dry Gas Lot No.: 302-402922402 Adjust. Gas Exp.: 12/14/2026
 0.300% Dry Gas Lot No.: 302-402757700 Adjust. Gas Exp.: 05/26/2026

Data Summary

Function	Result %BAC	Time hh:mm:ss	Barometric Pres. [hPa]	Comment(s) or Status Code
Ambient Air Blank	0.000	10:20:48		*TEST PASSED*
Control .04 Test 1 EC	0.039	10:21:26	995	*TEST PASSED*
Control .04 Test 1 IR	0.039	10:21:26	995	*TEST PASSED*
Ambient Air Blank	0.000	10:22:35		*TEST PASSED*
Control .04 Test 2 EC	0.039	10:22:51	995	*TEST PASSED*
Control .04 Test 2 IR	0.040	10:22:51	995	*TEST PASSED*
Ambient Air Blank	0.000	10:24:30		*TEST PASSED*
Control .08 Test 3 EC	0.078	10:25:06	995	*TEST PASSED*
Control .08 Test 3 IR	0.079	10:25:06	995	*TEST PASSED*
Ambient Air Blank	0.000	10:26:20		*TEST PASSED*
Control .08 Test 4 EC	0.080	10:26:34	995	*TEST PASSED*
Control .08 Test 4 IR	0.080	10:26:34	995	*TEST PASSED*
Ambient Air Blank	0.000	10:28:27		*TEST PASSED*
Control .16 Test 5 EC	0.157	10:29:00	995	*TEST PASSED*
Control .16 Test 5 IR	0.158	10:29:00	995	*TEST PASSED*
Ambient Air Blank	0.000	10:30:21		*TEST PASSED*
Control .16 Test 6 EC	0.160	10:30:33	995	*TEST PASSED*
Control .16 Test 6 IR	0.159	10:30:33	995	*TEST PASSED*
Ambient Air Blank	0.000	10:36:55		*TEST PASSED*
Control .30 Test 7 EC	0.301	10:37:30	995	*TEST PASSED*
Control .30 Test 7 IR	0.303	10:37:30	995	*TEST PASSED*
Ambient Air Blank	0.000	10:38:57		*TEST PASSED*
Control .30 Test 8 EC	0.307	10:39:10	995	*TEST PASSED*
Control .30 Test 8 IR	0.306	10:39:10	995	*TEST PASSED*
Ambient Air Blank	0.000	10:39:48		*TEST PASSED*

Result

All tests within acceptable tolerance.

Coordinator

Last Name: Fillimon - First Name: James MI: G Badge No.: 8306

On this date, I certified the above instrument in accordance with the Alcotest 9510 operator training and procedures established by the NJSP Office of Forensic Sciences.

Trp from Fillimon # 8306

Signed:

Date: 01/15/2026

ID: 2673

ALCOTEST 9510 CYLINDER INSTALLATION REPORT - INLET 1
South Brunswick
SERIAL NUMBER: ARMK-0097

Equipment

Inst. Model No.: ALCOTEST 9510 Serial No.: ARMK-0097
 Firmware: 8326739 1.5 Config.: 8326737 3.10 WinCE: 8326738 2.9
 Cyl1 Install File No.: 52 Cyl1 Install Date: 02/12/2025 Cyl1 Install No.: 2

Control Tests (0.100%)

Installation Inlet: #1 (Upper) Post test active Cyl.: #2 (Lower)
 Dry Gas Lot No.: 302-403034219 Dry Gas Lot Exp.: 05/01/2027

Data Summary

Function	Result %BAC	Time hh:mm:ss	Barometric Pres. [hPa]	Comment(s) or Status Code
Ambient Air Blank	0.000	09:33:37		*TEST PASSED*
Control Test 1			1021	*TEST PASSED*
EC Result	0.097	09:34:23		*TEST PASSED*
IR Result	0.100	09:34:23		*TEST PASSED*
Ambient Air Blank	0.000	09:35:40		*TEST PASSED*
Control Test 2			1021	*TEST PASSED*
EC Result	0.100	09:36:04		*TEST PASSED*
IR Result	0.101	09:36:04		*TEST PASSED*
Ambient Air Blank	0.000	09:37:22		*TEST PASSED*
Control Test 3			1021	*TEST PASSED*
EC Result	0.101	09:37:46		*TEST PASSED*
IR Result	0.101	09:37:46		*TEST PASSED*
Ambient Air Blank	0.000	09:38:25		*TEST PASSED*

Result

All tests within acceptable tolerance.

Coordinator

Last Name: Waldrop - First Name: Robert MI: W Badge No.: 8256

On this date, I certified the above instrument in accordance with the Alcotest 9510 operator training and procedures established by the NJSP Office of Forensic Sciences.

TR. Waldrop-8256

Signed:

Date: 02/12/2025

ID: 52

ALCOTEST 9510 CYLINDER INSTALLATION REPORT - INLET 2
South Brunswick
SERIAL NUMBER: ARMK-0097

Equipment

Inst. Model No.: ALCOTEST 9510 Serial No.: ARMK-0097
 Firmware: 8326739 1.5 Config.: 8326737 3.10 WinCE: 8326738 2.9
 Cyl2 Install File No.: 99 Cyl2 Install Date: 10/13/2025 Cyl2 Install No.: 2

Control Tests (0.100%)

Installation Inlet: #2 (Lower) Post test active Cyl.: #1 (Upper)
 Dry Gas Lot No.: 302-403378678 Dry Gas Lot Exp.: 06/20/2028

Data Summary

Function	Result %BAC	Time hh:mm:ss	Barometric Pres. [hPa]	Comment(s) or Status Code
Ambient Air Blank	0.000	08:51:25		*TEST PASSED*
Control Test 1			1009	*TEST PASSED*
EC Result	0.099	08:52:11		*TEST PASSED*
IR Result	0.101	08:52:11		*TEST PASSED*
Ambient Air Blank	0.000	08:53:24		*TEST PASSED*
Control Test 2			1009	*TEST PASSED*
EC Result	0.101	08:53:48		*TEST PASSED*
IR Result	0.102	08:53:48		*TEST PASSED*
Ambient Air Blank	0.000	08:55:01		*TEST PASSED*
Control Test 3			1009	*TEST PASSED*
EC Result	0.102	08:55:25		*TEST PASSED*
IR Result	0.102	08:55:25		*TEST PASSED*
Ambient Air Blank	0.000	08:55:58		*TEST PASSED*

Result

All tests within acceptable tolerance.

Coordinator

Last Name: RUBBE - First Name: ANTHONY MI: P Badge No.: 8352

On this date, I certified the above instrument in accordance with the Alcotest 9510 operator training and procedures established by the NJSP Office of Forensic Sciences.

TOR A. R. M. 8352

Signed: _____ Date: 10/13/2025 ID: 3018

CERTIFICATE OF ANALYSIS

EBS - ETHANOL BREATH STANDARD

Part Number: 4401036
DRAEGER MEDICAL SYSTEMS INC

Sales order: 1130435101
 Date: May 28, 2024

METHOD OF ANALYSIS: IR Breath Alcohol Analyzer
 ANALYTICAL ACCURACY: +/-0.002 BrAC or +/-2% whichever is greater.
 CALGAZ LOT#: 302-403034219
 ETHANOL IN NITROGEN

Manufactured Date: May 01, 2024
 Product Expiration: May 01, 2027

COMPONENT	PPM	(BrAC)
ETHANOL	260.5PPM	(0.100)
NITROGEN	BAL	
AVERAGE ANALYTICAL VALUE	PPM	(BrAC)
ETHANOL	263.0	(0.101)

REFERENCE STANDARD	CYLINDER	CONCENTRATION PPM
N.M.I. TRACEABLE STANDARDS*	ND38424	260.7

* CERTIFICATION TRACEABLE TO NATIONAL METROLOGY INSTITUTE TRACEABLE STANDARDS

TRACEABILITY

Preparation:

Gas mixtures manufactured with balances calibrated by an ISO 17025 accredited company using NIST traceable weights and meets or exceeds the requirements of NIST Handbook 44.

Traceable certificate numbers 3445312 and 3398673.

Analytical:


Analytical Instruments Calibrated Using NMI Traceable Standards.

Certification Numbers: A679-20190918, D049803-20220329

No effecting environmental conditions during analysis.

*NMI is recognized by NIST through the Mutual Recognition Agreement (CIPM MRA).

CALGAZ calibration devices were found to meet all applicable requirements of the National Highway Traffic Safety Administration Model Specifications for calibrating units for breath alcohol testers.

APPROVED BY: 

*We certify that all the cylinders for the Lot numbers identified herein are manufactured and tested within the requirements of CFR 49 part 178.65 and that physical and chemical test reports are on file and copies will be furnished upon request.

CALGAZ, a division of Airgas USA LLC
 821 Chesapeake Drive, Cambridge, MD 21613-0149
 Phone: (410) 228-6400 Fax: (410) 228-4251

CERTIFICATE OF ANALYSIS

EBS - ETHANOL BREATH STANDARD

Part Number: 4401036

DRAEGER MEDICAL SYSTEMS INC

Sales order: 1141350717

Date: July 25, 2025

METHOD OF ANALYSIS: IR Breath Alcohol Analyzer

ANALYTICAL ACCURACY: +/-0.002 BrAC or +/-2% whichever is greater.

CALGAZ LOT#: 302-403376678

ETHANOL IN NITROGEN

Manufactured Date: June 20, 2025

Product Expiration: June 20, 2028

COMPONENT	PPM	(BrAC)
ETHANOL	260.5PPM	(0.100)
NITROGEN	BAL	
AVERAGE ANALYTICAL VALUE	PPM	(BrAC)
ETHANOL	262.6	(0.101)
REFERENCE STANDARD	CYLINDER	CONCENTRATION PPM
N.M.I. TRACEABLE STANDARDS*	ND49826	260.1

* CERTIFICATION TRACEABLE TO NATIONAL METROLOGY INSTITUTE TRACEABLE STANDARDS

TRACEABILITY

Preparation:

Gas mixtures manufactured with balances calibrated by an ISO 17025 accredited company using NIST traceable weights and meets or exceeds the requirements of NIST Handbook 44. Traceable certificate numbers 3445312 and 3398673.

Analytical:

Analytical Instruments Calibrated Using NMI Traceable Standards.


Certification Numbers: A679-20190918, D049803-20220329

No effecting environmental conditions during analysis.

*NMI is recognized by NIST through the Mutual Recognition Agreement (CIPM MRA).

CALGAZ calibration devices were found to meet all applicable requirements of the National Highway Traffic Safety Administration Model Specifications for calibrating units for breath alcohol testers.

APPROVED BY:




"We certify that all the cylinders for the Lot numbers identified herein are manufactured and tested within the requirements of CFR 49 part 178.65 and that physical and chemical test reports are on file and copies will be furnished upon request."

CALGAZ, a division of Airgas USA LLC

821 Chesapeake Drive, Cambridge, MD 21613-0149

Phone: (410) 228-6400

Fax: (410) 228-4251



Dräger

Alcotest 9510

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest 9510 has been tested for accuracy and found to be in compliance with the National Highway Traffic Safety Administration Standard for evidential breath testing devices. The Alcotest MKIII-C is compliant as a "mobile" and "nonmobile" EBT with 49 FR 48854, 49 FR 48864, and 58 FR 48705. The manufacturer recommends accuracy verification of this instrument within 12 months of the calibration date below, or sooner, according to your state's specifications.

Certification Date:

Serial Number:

11/28/2023

ARMK-0097

Draeger, Inc.

J MB



State of New Jersey

OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
POST OFFICE BOX 7068
WEST TRENTON, NJ 08628-0068
(609) 882-2000

PHILIP D. MURPHY
Governor

TAHESHA L. WAY
Lt. Governor

MATTHEW J. PLATEKIN
Attorney General

COLONEL PATRICK J. CALLAHAN
Superintendent

CERTIFICATION OF ANALYSIS 0.100 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.1174 to 0.1246 grams per 100 milliliters of solution.

MANUFACTURER: Draeger, Inc.

ANALYSIS DATE: 07/18/2024

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 24210

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1195 to 0.1217 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is June 11, 2026.

As OFS Director for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Michael Kennedy
Michael Kennedy
Director
NJSP Office of Forensic Sciences

Sworn to and subscribed before me this 17 day of July, 2024.

Notary

KAREN E. STAHL
NOTARY PUBLIC OF NEW JERSEY
Commission # 50110522
My Commission Expires 8/13/2024

KAREN E. STAHL
NOTARY PUBLIC OF NEW JERSEY
Commission # 50110522

KAREN E. STAHL
NOTARY PUBLIC OF NEW JERSEY
Commission # 50110522

KAREN E. STAHL
NOTARY PUBLIC OF NEW JERSEY
Commission # 50110522



"An Internationally Accredited Agency"

New Jersey Is An Equal Opportunity Employer
Printed on Recycled Paper and Recyclable



Customer: DRAEGER INC
7256 S SAM HOUSTON PKWY W
STE 100
HOUSTON TX 77085
PO Number: SUC4303700862



Certificate/ISO Number: 5-F8B2G-280-1 Revision 0

Manufacturer: Draeger Safety AG & Co. KGaA
Model Number: X-Cal 2000
Description: Breath Alcohol Simulator
Serial Number: ARND-0002
ID: NONE

As-Found: Out Of Tolerance
As-Left: In Tolerance

Issue Date: Aug 20, 2025
Calibration Date: Aug 20, 2025
Due Date: Aug 20, 2026

Calibrated To: Manufacturer Specification
Calibration Procedure: 1-AC103519-2

Transcal Calibration Laboratories have been audited and found in compliance with ISO/IEC 17025:2017. Accredited calibrations performed within the Lab Scope of Accreditation are indicated by the presence of the Accrediting Body Logo and Certificate Number. Any measurements on an accredited calibration not covered by the Lab Scope of Accreditation are listed in the notes section of the certificate. SCC, NRC, CLAS or ANAB do not guarantee the accuracy of an individual calibration by accredited laboratories.

Transcal calibrations, as applicable, are performed in compliance with the requirements of the Transcal Quality Manual QAC-901-009, the customer Purchase Order and/or Quality Agreement requirements, ISO 9001:2015, ANSI/NCSL Z540.1-1994 (R2002), and ISO 10012:2003, as applicable. When specified contractually, the requirements of ISO TS 16949:2009, IEC 61331, 19CFR80 App. B, ASME NQA-1:2012, and ANSI/NCSL Z540.3-2006 (R2013) are also covered.

Complete records of work performed are maintained by Transcal and are available for inspection. Laboratory standards used in the performance of this calibration are listed on this certificate.

Transcal documents the traceability of measurements to the SI units through the National Institute of Standards and Technology (NIST), or the National Research Council of Canada (NRC), or other national measurement institutes (NMI) that are signatories to the OIML Mutual Recognition Arrangement, or accepted fundamental and/or related physical constants, or by the use of specified methods, consensus standards or ratio type measurements. Documentation supporting traceability information is available for review upon written request at a Transcal facility. The measured quantity and the measurement uncertainty are required for further dissemination of traceability.

Uncertainties are reported with a coverage factor k=2, providing a level of confidence of approximately 95%. All calibrations have been performed using processes having a TUR of 4:1 or better (4:1 for mass calibrations), unless otherwise noted. The Test Uncertainty Ratio (TUR) is calculated in accordance with NCSL International RP-18. For mass calibrations, conventional mass references to 0.0 g/cm³.


The results in this report relate only to the item calibrated or tested. Recorded calibration data is valid at the time of calibration within the stated uncertainties at the environmental conditions noted. The determination of compliance to the specification is specific to the model/serial no./ID no. referenced above based on the tolerances shown; these tolerances are either the original equipment manufacturer's (OEM's) warranted specifications or the client's requested specifications. Any number of factors can cause a unit to drift out of tolerance at any time following its calibration. Limitations on the uses of this instrument are detailed in the OEM's operating instructions. This certificate may not be reproduced except in full, without the written approval of Transcal. Additional information, if applicable may be included on separate report(s).

Notes:

Unit was received Out of Tolerance and adjustments were made for best overall accuracy.

The OOT readings were verified.

Customer: DRAEGER INC
7256 S.SAM HOUSTON PKWY W
STE 100
HOUSTON, TX 77085
PO Number: SUC4303700862



Certificate/SO Number: 5-F8B2G-280-1 Revision 0

As Found Data

Description	Setpoints	Accuracy	Low Limit	High Limit	As Found	O T	Cal Process Uncertainty (k=2; s)	Measurement Uncertainty (k=2; s)	Units	TUR
Function Checks										
Bubble Check			P	P	P					
Seal Check			P	P	P					
Temperature Source: Accuracy Test										
Accuracy Test	34.00°C	±(0.02 °C)	33.98	34.02	33.97 °C	*	1.56-002	1.86-002	°C	1.3:1
Temperature Source: Stability Test										
Stability Test	0.00°C	±(0.02 °C)	-0.02	0.02	0.00 °C		1.56-002	1.86-002	°C	1.3:1

As Left Data

Description	Setpoints	Accuracy	Low Limit	High Limit	As Left	O T	Cal Process Uncertainty (k=2; s)	Measurement Uncertainty (k=2; s)	Units	TUR
Function Checks										
Bubble Check			P	P	P					
Seal Check			P	P	P					
Temperature Source: Accuracy Test										
Accuracy Test	34.00°C	±(0.02 °C)	33.98	34.02	34.00 °C		1.56-002	1.66-002	°C	1.3:1
Temperature Source: Stability Test										
Stability Test	0.00°C	±(0.02 °C)	-0.02	0.02	0.00 °C		1.56-002	1.66-002	°C	1.3:1

Field not applicable.

Customer: DRAEGER INC.

7256 S SAM HOUSTON PKWY W.

STE 100

HOUSTON, TX 77085

PO Number: SUC4303700862



Certificate/SO Number: 5-F8B2G-280-1 Revision 0

Traceable Standards

Asset	Manufacturer	Model Number	Description	Cal Date	Due Date	Traceability Number	Use
05H4479	AccuMac Corporation	AM1760-12-S	Secondary SPRT	12-Aug-24	31-Aug-25	H32XM-4-1	AF/AL
HP927312	Hart Scientific/Fluke	1675	Super Thermometer	10-Jul-24	31-Jan-26	5-8-HP927312-9-1	AF/AL

The use of the standard is defined as: AF - used for as-found readings, AL - used for as-left readings.


Environmental Data

Temperature	Relative Humidity	Temp / RH Asset	Lab Area	Lab Description
69.70°F / 20.94°C	56.60%	DewK11	G	Temperature

Decision Rule

When compliance statements are present, they are reported without factoring in the effects of uncertainty and comply with the guidelines as follows: The acceptance zone is defined as: less than or equal to the high limit, and/or greater than or equal to the low limit. The rejection zones are defined as greater than the high limit and/or less than the low limit. Single measurement results in the acceptance zone are identified as in-tolerance. Single measurement results in the rejection zone are identified as out-of-tolerance (OOT). When all measurement results are in the acceptance zone for repeated measurements, for the same characteristic, the test is identified as in-tolerance. For repeated characteristic measurements, a single measurement result in the rejection zone, will cause the test to be identified as out-of-tolerance (OOT). Data rejection for cause, (outliers) is permitted after the "Determining and Verifying Out Of Tolerance (OOT) and/or Op Fail Readings" procedure outlined in this document has been completed and the anomalous reading cannot be repeated, and the anomalous reading does not represent the system under test. Statements of conformity are binary.

Customer: DRAEGER INC
7256 S SAM HOUSTON PKWY W
STE 100
HOUSTON, TX 77085
PO Number: SUC4303700852



Certificate/SO Number: 5-F8B2G-280-1 Revision 0


Legend

Topic	Description
Accuracy	UUT specification that establishes expected tolerances and a time limit (calibration interval) over which the instrument is expected to hold these tolerances
As Found	Initial measurement results
As Left	Measurement results after adjustment and/or repair
Blank Data Field	Test is not applicable for the UUT
Cal Process Uncertainty (CPU)	The uncertainty of calibration process for the reported measurement result
Calibration Date	Indicates the date that the calibration was completed
Cover Factor (k)	A measure of uncertainty that defines an interval about the measurement result
Due Date	Indicates the end of the calibration cycle as requested by the customer
Issue Date	Indicates the date that the calibration has passed the Data Review Process and was signed by an authorized signatory or the date that a revision to the original certificate has been issued
Low / High Limits	Establishes UUT acceptable performance limits for the test measurement
Measurement Uncertainty	The dispersion of the values attributed to a measured quantity
CPA	Out of Acceptance (#)
OOT	Out of Tolerance (%)
Setpoints	Measurement target values
Traceability	Unbroken chain of comparisons relating an instrument's measurements to a known standard(s)
Traceability Number	Unique identifier(s) used to document traceability of calibration standards
TUR	Test Uncertainty Ratio, ratio of the tolerance or specification of the test measurement in relation to the uncertainty in measurement results
UUT	Unit Under test

CALIBRATED
BY **TRANSAT**

CERTIFICATE OF CALIBRATION

Customer: DRAEGER INC
7256 S SAM HOUSTON PKWY W
STE 100
HOUSTON, TX 77065
PO Number: SUC4303700882



Certificate/SO Number: 5-F8B2G-280-1 Revision 0

Calibrated At:
18118 Park Row
Houston, TX 77064



Date Received: August 08, 2025
Service Level: R9

Facility Responsibility:
18118 Park Row
Houston, TX 77064
800.426.1470

Calibrated By:

Jose Martinez
Calibration Technician

Electronically Signed By:
Jose Martinez

Aug 20, 2025
11:56:02 -04:00

Reviewed By:

Josh Soltau
Lab Manager


Electronically Signed By:
Luis Arac for

Aug 20, 2025
13:57:50 -04:00

Certificate - Page 5 of 5
Reprinted on August 26, 2025

Customer Number: 1-859111-000
CPS: F20-014R11-072723 FP001R9-4/9/2021

Customer: DRAEGER INC
7258 S SAM HOUSTON PKWY W
STE 100
HOUSTON, TX 77065
PO Number: SUC4308700862



Certificate/ISO Number: 5-F8B2G-100-1 Revision 0

Manufacturer: Wika Instr/Mensor Corp/Trend
Model Number: CPG2300
Description: Portable Barometer
Serial Number: 41001272
ID: NONE

As-Found: Out Of Tolerance
As-Left: In Tolerance

Issue Date: Aug 25, 2025
Calibration Date: Aug 25, 2025
Due Date: Aug 25, 2026

Calibrated To: Manufacturer Specification
Calibration Procedure: 1-AC10728B-0

Transcat Calibration Laboratories have been audited and found in compliance with ISO/IEC 17025:2017. Accredited calibrations performed within the Lab Scope of Accreditation are indicated by the presence of the Accrediting Body Logo and Certificate Number. Any measurements on an accredited calibration not covered by the Lab Scope of Accreditation are listed in the notes section of the certificate. SCC, NRC, CLAS or ANAB do not guarantee the accuracy of an individual calibration by accredited laboratories.

Transcat calibrations, as applicable, are performed in compliance with the requirements of the Transcat Quality Manual QAC-PQI-000, the customer Purchase Order and/or Quality Agreement requirements, ISO 9001:2015, ANSI/INCISL Z540.1-1994 (R2002), and ISO 10012:2003, as applicable, when specified contractually, the requirements of ISO TS16949:2009, 10CFR21, 10CFR50, App. B, ASME NQA-1(2012), and ANSI/INCISL Z540.5-2006 (R2013) are also covered.

Complete records of work performed are maintained by Transcat and are available for inspection. Laboratory standards used in the performance of this calibration are listed on this certificate.

Transcat documents the traceability of measurements to the SI units through the National Institute of Standards and Technology (NIST), or the National Research Council of Canada (NRC), or other national measurement institutes (NMI) that are signatories to the CIPM Mutual Recognition Arrangement, or accepted fundamental and/or natural physical constants, or by the use of specified methods, consensus standards or ratio type measurements. Documentation supporting traceability information is available for review upon written request at a Transcat facility. The measured quantity and the measurement uncertainty are required for further dissemination of traceability.


Uncertainties are reported with a coverage factor $k=2$, providing a level of confidence of approximately 95%. All calibrations have been performed using processes having a TUR of 4:1 or better (3:1 for mass calibrations), unless otherwise noted. The Test Uncertainty Ratio (TUR) is calculated in accordance with NCSL International RP-18. For mass calibrations, conventional mass referenced to 6.0 g/cm³.

The results in this report relate only to the item calibrated or tested. Recorded calibration data is valid at the time of calibration within the stated uncertainties at the environmental conditions noted. The determination of compliance to the specification is specific to the model/serial no./ID no. referenced above based on the tolerances shown, these tolerances are either the original equipment manufacturer (OEM's) warranted specifications or the client's requested specifications. Any number of factors can cause a unit to drift out of tolerance at any time following its calibration. Limitations on the uses of this instrument are detailed in the OEM's operating instructions. This certificate may not be reproduced except in full, without the written approval of Transcat. Additional information, if applicable may be included on separate report(s).

Notes:

Received out of tolerance, Calibrated, Returned in tolerance. The QOT readings were verified.

Customer: DRAEGER INC
7256 S SAM HOUSTON PKWY W
STE 100
HOUSTON, TX 77065
PO Number: SUC4303700862




Certificate/SO Number: 5-F8B2G-100-1 Revision 0

As Found Data

Description	Setpoints	Accuracy	As Found Data			Cal Process Uncertainty (k=2,1.5)	Measurement Uncertainty (k=2,1.5)	Units	TUR
			Low Limit	High Limit	As Found				
Pressure Measure: 552 to 1172 mbara Range									
550.1 mbara	±(0.015% FS)	549.9	550.3	549.7 mbara	9.4e-003	6.1e-002	mbara	21.4 : 1	
610.0 mbara	±(0.015% FS)	609.6	610.2	609.6 mbara	1.0e-002	5.9e-002	mbara	19.3 : 1	
680.4 mbara	±(0.015% FS)	680.2	680.5	680.0 mbara	1.2e-002	5.9e-002	mbara	17.3 : 1	
734.3 mbara	±(0.015% FS)	734.1	734.5	733.9 mbara	1.2e-002	5.9e-002	mbara	16.0 : 1	
804.6 mbara	±(0.015% FS)	804.4	804.8	804.3 mbara	1.4e-002	6.0e-002	mbara	14.6 : 1	
864.9 mbara	±(0.015% FS)	864.7	865.1	864.6 mbara	1.5e-002	6.0e-002	mbara	13.6 : 1	
924.9 mbara	±(0.015% FS)	924.7	925.1	924.6 mbara	1.6e-002	6.0e-002	mbara	12.7 : 1	
985.2 mbara	±(0.015% FS)	985.0	985.4	984.9 mbara	1.7e-002	6.0e-002	mbara	11.9 : 1	
1043.9 mbara	±(0.015% FS)	1043.7	1044.1	1043.5 mbara	1.8e-002	6.1e-002	mbara	11.3 : 1	
1114.2 mbara	±(0.015% FS)	1114.0	1114.4	1113.9 mbara	1.9e-002	6.1e-002	mbara	10.6 : 1	
1174.6 mbara	±(0.015% FS)	1174.4	1174.8	1174.2 mbara	2.0e-002	6.1e-002	mbara	10.0 : 1	
924.9 mbara	±(0.015% FS)	924.7	925.1	924.6 mbara	1.6e-002	6.0e-002	mbara	12.7 : 1	
864.9 mbara	±(0.015% FS)	864.7	865.1	864.6 mbara	1.5e-002	6.0e-002	mbara	13.6 : 1	
804.6 mbara	±(0.015% FS)	804.4	804.8	804.3 mbara	1.4e-002	6.0e-002	mbara	14.6 : 1	

Customer: DRAEGER INC
 7266 S SAM HOUSTON PKWY W
 STE 100
 HOUSTON, TX 77085
 PO Number: SUC4303700862



Certificate/ISO Number: 5-F8B2G-100-1 Revision 0

As Left Data

Description	Setpoints	Accuracy	Low Limit	High Limit	As Left	Cal Process			Units	TUR
						O	U	T		
Pressure Measure: 552 to 1172 mbara Range										
	550.1 mbara	±(0.015% FS)	549.9	550.3	550.1 mbara	9.4e-003	6.1e-002	6.1e-002	mbara	21.4 : 1
	610.0 mbara	±(0.015% FS)	609.8	610.2	610.0 mbara	1.0e-002	5.9e-002	5.9e-002	mbara	18.3 : 1
	680.4 mbara	±(0.015% FS)	680.2	680.6	680.4 mbara	1.2e-002	5.9e-002	5.9e-002	mbara	17.3 : 1
	734.3 mbara	±(0.015% FS)	734.1	734.5	734.3 mbara	1.2e-002	5.9e-002	5.9e-002	mbara	16.0 : 1
	804.6 mbara	±(0.015% FS)	804.4	804.8	804.7 mbara	1.4e-002	6.0e-002	6.0e-002	mbara	14.6 : 1
	864.9 mbara	±(0.015% FS)	864.7	865.1	865.0 mbara	1.5e-002	6.0e-002	6.0e-002	mbara	13.6 : 1
	924.9 mbara	±(0.015% FS)	924.7	925.1	925.0 mbara	1.8e-002	6.0e-002	6.0e-002	mbara	12.7 : 1
	985.2 mbara	±(0.015% FS)	985.0	985.4	985.3 mbara	1.7e-002	6.0e-002	6.0e-002	mbara	11.9 : 1
	1043.8 mbara	±(0.015% FS)	1043.6	1044.0	1043.9 mbara	1.8e-002	6.1e-002	6.1e-002	mbara	11.3 : 1
	1114.2 mbara	±(0.015% FS)	1114.0	1114.4	1114.3 mbara	1.9e-002	6.1e-002	6.1e-002	mbara	10.6 : 1
	1174.6 mbara	±(0.015% FS)	1174.4	1174.8	1174.6 mbara	2.0e-002	6.1e-002	6.1e-002	mbara	10.0 : 1
	924.9 mbara	±(0.015% FS)	924.7	925.1	925.0 mbara	1.6e-002	6.0e-002	6.0e-002	mbara	12.7 : 1
	864.9 mbara	±(0.015% FS)	864.7	865.1	865.0 mbara	1.5e-002	6.0e-002	6.0e-002	mbara	13.6 : 1
	804.6 mbara	±(0.015% FS)	804.4	804.8	804.7 mbara	1.4e-002	6.0e-002	6.0e-002	mbara	14.6 : 1

Field not applicable.

Traceable Standards


Asset	Manufacturer	Model Number	Description	Cal Date	Due Date	Traceability Number	Use
DW11BA	Fluke/DH Instruments	PG7601	Piston Gauge	31-Jul-25	31-Jul-26	5-4DW11BA-20-1	AF/AL
DW11CA	Fluke/DH Instruments	MS-AMH-38	AMH Mass Set	5-Jun-25	30-Sep-25	5-4DW11CA-40-1	AF/AL
DW11LOW	Fluke/DH Instruments	PC-7100/7600-10-TC	Gas Piston-Cylinder Module	9-Apr-22	30-Apr-27	5-4DW11LOW-3-1	AF/AL
DW11MASS	Fluke/DH Instruments	MS-AMH-38	AMH Mass Set	5-Mar-25	31-Mar-26	5-4DW11MASS-12-1	AF/AL

The use of the standard is defined as: AF - used for as-found readings, AL - used for as-left readings.

Customer: DRAEGER INC

7256 S SAM HOUSTON PKWY W
STE 100
HOUSTON, TX 77085

PO Number: SUC4303700892



Certificate/SO Number: 5-F8B2G-100-1 Revision 0

Environmental Data


	Temperature	Relative Humidity	Temp / RH Asset		Lab Area	Lab Description
			DewK10	DewK10		
As Found:	69.31°F / 20.73°C	49.60%	DewK10	B		GP Pressure
As Left:	69.27°F / 20.71°C	52.20%	DewK10	B		GP Pressure

Decision Rule

When compliance statements are present, they are reported without factoring in the effects of uncertainty and comply with the guidelines as follows: The acceptance zone is defined as: less than or equal to the high limit, and/or greater than or equal to the low limit. The rejection zones are defined as greater than the high limit and/or less than the low limit. Single measurement results in the acceptance zone are identified as in-tolerance. Single measurement results in the rejection zone are identified as out-of-tolerance (OOT). When all measurement results are in the acceptance zone for repeated measurements, for the same characteristic, the test is identified as in-tolerance. For repeated characteristic measurements, a single measurement result in the rejection zone will cause the test to be identified as out-of-tolerance (OOT). Data rejection for cause (outliers) is permitted after the "Determining and Verifying Out Of Tolerance (OOT) and/or Op Fail Readings" procedure outlined in this document has been completed and the anomalous reading cannot be repeated, and the anomalous reading does not represent the system under test. Statements of conformity are binary.

CALIBRATED BY TRANSQIT

CERTIFICATE OF CALIBRATION



Customer: DRAEGER INC
7256 S SAM HOUSTON PKWY W
STE 100
HOUSTON, TX 77085
PO Number: SUC4303700862

Certificate/SO Number: 5-F8B2G-100-1 Revision 0


Legend

Topic	Description
Accuracy	UUT specification that establishes expected tolerances and a time limit (calibration interval) over which the instrument is expected to hold these tolerances
As Found	Initial measurement results
As Left	Measurement results after adjustment and/or repair
Blank Data Field	Test is not applicable for the UUT
Cal Process Uncertainty (CPU)	The uncertainty of calibration process for the reported measurement result
Calibration Date	Indicates the date that the calibration was completed
Cover Factor (k)	A measure of uncertainty that defines an interval about the measurement result
Due Date	Indicates the end of the calibration cycle as requested by the customer
Issue Date	Indicates the date that the calibration has passed the Data Review Process and was signed by an authorized signatory or the date that a revision to the original certificate has been issued
Low / High Limits	Establishes UUT acceptable performance limits for the test measurement
Measurement Uncertainty	The dispersion of the values attributed to a measured quantity
OOA	Out of Acceptance (#)
OOT	Out of Tolerance (%)
Setpoints	Measurement target values
Traceability	Unbroken chain of comparisons relating an instrument's measurements to a known standard(s)
Traceability Number	Unique identifier(s) used to document traceability of calibration standards
TUR	Test Uncertainty Ratio, ratio of the tolerance or specification of the test measurement in relation to the uncertainty in measurement results
UUT	Unit Under Test

CALIBRATED
BY **TRANSIT**

CERTIFICATE OF CALIBRATION

Customer: DRAEGER INC
7256 S SAM HOUSTON PKWY W
SITE 100
HOUSTON, TX 77085
PO Number: SUC4303700862



Certificate/SO Number: 5-F8B2G-100-1 Revision 0

Calibrated At:
16115 Park Row
Houston, TX 77084

Facility Response/Label:
16115 Park Row
Houston, TX 77084
800-828-1470



Date Received: August 08, 2025
Service Level: R9

Calibrated By:

Electronically Signed By:
Evan Copeland
Evan Copeland
Calibration Technician

Aug 25, 2025
19:31:27 -0400

Reviewed By:

Electronically Signed By:
Graham Walker for
Graham Walker for

Aug 25, 2025
15:24:39 -0400

Certificate - Page 6 of 6
Reprinted on August 27, 2025

Customer Number: 1-659111-000
OFS-F20-014R11 07/27/23 FP001RS #19/2021

CERTIFICATE OF ANALYSIS

EBS - ETHANOL BREATH STANDARD

Part Number: 4401036

DRAEGER MEDICAL SYSTEMS INC

Sales order: 1130434779

Date: May 23, 2024

METHOD OF ANALYSIS: IR Breath Alcohol Analyzer

ANALYTICAL ACCURACY: +/-0.002 BrAC or +/-2% whichever is greater.

CALGAZ LOT#: 302-403034216

Manufactured Date: April 30, 2024

ETHANOL IN NITROGEN

Product Expiration: April 30, 2027

COMPONENT	PPM	(BrAC)
ETHANOL	260.5PPM	(0.100)
NITROGEN	BAL	
AVERAGE ANALYTICAL VALUE	PPM	(BrAC)
ETHANOL	264.1	(0.101)
REFERENCE STANDARD	CYLINDER	CONCENTRATION PPM
N.M.I. TRACEABLE STANDARDS*	ND38424	260.7

* CERTIFICATION TRACEABLE TO NATIONAL METROLOGY INSTITUTE TRACEABLE STANDARDS

TRACEABILITY

Preparation:

Gas mixtures manufactured with balances calibrated by an ISO 17025 accredited company using NIST traceable weights and meets or exceeds the requirements of NIST Handbook 44.

Traceable certificate numbers 3445312 and 3398673.

Analytical:

Analytical Instruments Calibrated Using NMI Traceable Standards.


Certification Numbers: A679-20190918, D049803-20220329

No effecting environmental conditions during analysis.

*NMI is recognized by NIST through the Mutual Recognition Agreement (CIPM MRA).

CALGAZ calibration devices were found to meet all applicable requirements of the National Highway Traffic Safety Administration Model Specifications for calibrating units for breath alcohol testers.

APPROVED BY:



"We certify that all the cylinders for the Lot numbers identified herein are manufactured and tested within the requirements of CFR 49 part 178.65 and that physical and chemical test reports are on file and copies will be furnished upon request."

CALGAZ, a division of Airgas USA LLC
 821 Chesapeake Drive, Cambridge, MD 21613-0149
 Phone: (410) 228-6400 Fax: (410) 228-4251

CERTIFICATE OF ANALYSIS

EBS - ETHANOL BREATH STANDARD

DEPT OF LAW AND PUBLIC SAFETY

Sales order: 1120656707

Date: May 30, 2023

METHOD OF ANALYSIS: IR Breath Alcohol Analyzer
ANALYTICAL ACCURACY: +/-0.002 BrAC or +/-2% whichever is greater
CALGAZ LOT#: 302-402730318
ETHANOL IN NITROGEN

Product Expiration: April 27, 2026

COMPONENT	PPM	(BrAC)
ETHANOL	104.2PPM	(0.040)
NITROGEN	BAL	
AVERAGE ANALYTICAL VALUE	PPM	(BrAC)
ETHANOL	107.8	(0.041)

REFERENCE STANDARD	CYLINDER	CONCENTRATION PPM
N.M.I. TRACEABLE STANDARDS*	ND38424	260.7

* CERTIFICATION TRACEABLE TO NATIONAL METROLOGY INSTITUTE TRACEABLE STANDARDS:

TRACEABILITY

Preparation:

Gas mixtures manufactured with balances calibrated by an ISO 17025 accredited company using NIST traceable weights and meets or exceeds the requirements of NIST Handbook 44.

Traceable certificate numbers 3445312 and 3398673.

Analytical:

Analytical Instruments Calibrated Using NMI Traceable Standards.

Certification Numbers: A679-20190918, D049803-20220329

No effecting environmental conditions during analysis.

*NMI is recognized by NIST through the Mutual Recognition Agreement (CIPM MRA).

CALGAZ calibration devices were found to meet all applicable requirements of the National Highway Traffic Safety Administration Model Specifications for calibrating units for breath alcohol testers.

Manufactured Date: April 27, 2023

APPROVED BY:



"We certify that all the cylinders for the Lot numbers identified herein are manufactured and tested within the requirements of CFR 49 part 178.65 and that physical and chemical test reports are on file and copies will be furnished upon request."

CALGAZ, a division of Airgas USA LLC

821 Chesapeake Drive, Cambridge, MD 21613-0149

Phone: (410) 228-6400 Fax: (410) 228-4251

CERTIFICATE OF ANALYSIS

EBS - ETHANOL BREATH STANDARD

DEPT OF LAW AND PUBLIC SAFETY

Sales order: 1120656618

Date: May 25, 2023

METHOD OF ANALYSIS: IR Breath Alcohol Analyzer
ANALYTICAL ACCURACY: +/-0.002 BrAC or +/-2% whichever is greater.
CALGAZ LOT#: 302-402732434
ETHANOL IN NITROGEN

Product Expiration: April 28, 2026

COMPONENT	PPM	(BrAC)
ETHANOL	208.4PPM	(0.080)
NITROGEN	BAL	
AVERAGE ANALYTICAL VALUE	PPM	(BrAC)
ETHANOL	210.4	(0.081)

REFERENCE STANDARD	CYLINDER	CONCENTRATION PPM
N.M.I. TRACEABLE STANDARDS*	ND38424	260.7

* CERTIFICATION TRACEABLE TO NATIONAL METROLOGY INSTITUTE TRACEABLE STANDARDS

TRACEABILITY

Preparation:

Gas mixtures manufactured with balances calibrated by an ISO 17025 accredited company using NIST traceable weights and meets or exceeds the requirements of NIST Handbook 44.

Traceable certificate numbers 3445312 and 3398673.

Analytical:

Analytical Instruments Calibrated Using NMI Traceable Standards.

Certification Numbers: A679-20190918, D049803-20220329


No effecting environmental conditions during analysis.

*NMI is recognized by NIST through the Mutual Recognition Agreement (CIPM MRA).

CALGAZ calibration devices were found to meet all applicable requirements of the National Highway Traffic Safety Administration Model Specifications for calibrating units for breath alcohol testers.

Manufactured Date: April 28, 2023

APPROVED BY:



"We certify that all the cylinders for the Lot numbers identified herein are manufactured and tested within the requirements of CFR 49 part 178.65 and that physical and chemical test reports are on file and copies will be furnished upon request."

CALGAZ, a division of Airgas USA LLC
821 Chesapeake Drive, Cambridge, MD 21613-0149
Phone: (410) 228-6400 Fax: (410) 228-4251

CERTIFICATE OF ANALYSIS

EBS - ETHANOL BREATH STANDARD

Part Number: 4401040NJ
DRAEGER MEDICAL SYSTEMS INC

Sales order: 1126209454
Date: December 18, 2023

METHOD OF ANALYSIS: IR Breath Alcohol Analyzer
ANALYTICAL ACCURACY: +/-0.002 BrAC or +/-2% whichever is greater.
CALGAZ LOT#: 302-402922402
ETHANOL IN NITROGEN

Product Expiration: December 14, 2026

COMPONENT	PPM	(BrAC)
ETHANOL	416.8PPM	(0.160)
NITROGEN	BAL	
AVERAGE ANALYTICAL VALUE	PPM	(BrAC)
ETHANOL	421.3	(0.162)

REFERENCE STANDARD	CYLINDER	CONCENTRATION PPM
N.M.I. TRACEABLE STANDARDS*	ND38424	260.7

* CERTIFICATION TRACEABLE TO NATIONAL METROLOGY INSTITUTE TRACEABLE STANDARDS

TRACEABILITY

Preparation:

Gas mixtures manufactured with balances calibrated by an ISO 17025 accredited company using NIST traceable weights and meets or exceeds the requirements of NIST Handbook 44.

Traceable certificate numbers 3445312 and 3398673.

Analytical:

Analytical Instruments Calibrated Using NMI Traceable Standards.

Certification Numbers: A679-20190918, D049803-20220329

No effecting environmental conditions during analysis.

*NMI is recognized by NIST through the Mutual Recognition Agreement (CIPM MRA).

CALGAZ calibration devices were found to meet all applicable requirements of the National Highway Traffic Safety Administration Model Specifications for calibrating units for breath alcohol testers.

Manufactured Date: December 14, 2023

APPROVED BY: _____



"We certify that all the cylinders for the Lot numbers identified herein are manufactured and tested within the requirements of CFR 49 part 178.65 and that physical and chemical test reports are on file and copies will be furnished upon request."

CALGAZ, a division of Airgas USA LLC
821 Chesapeake Drive, Cambridge, MD 21613-0149
Phone: (410) 228-6400 Fax: (410) 228-4251

CERTIFICATE OF ANALYSIS

EBS - ETHANOL BREATH STANDARD

DEPT OF LAW AND PUBLIC SAFETY

Sales order: 120656632
Date: May 31, 2023

METHOD OF ANALYSIS: IR Breath Alcohol Analyzer
ANALYTICAL ACCURACY: +/-0.002 BrAC or +/-2% whichever is greater.
CALGAZ LOT#: 302-402757700
ETHANOL IN NITROGEN

Product Expiration: May 26, 2026

COMPONENT	PPM	(BrAC)
ETHANOL	781.5PPM	(0.300)
NITROGEN	BAL	
AVERAGE ANALYTICAL VALUE		
	PPM	(BrAC)
ETHANOL	794.4	(0.305)

REFERENCE STANDARD	CYLINDER	CONCENTRATION PPM
N.M.I. TRACEABLE STANDARDS*	ND38424	260.7

* CERTIFICATION TRACEABLE TO NATIONAL METROLOGY INSTITUTE TRACEABLE STANDARDS

TRACEABILITY

Preparation:

Gas mixtures manufactured with balances calibrated by an ISO 17025 accredited company using NIST traceable weights and meets or exceeds the requirements of NIST Handbook 44.
Traceable certificate numbers 3445312 and 3398673.

Analytical:


Analytical Instruments Calibrated Using NMI Traceable Standards.
Certification Numbers: A679-20190918, D049803-20220329

No effecting environmental conditions during analysis.

*NMI is recognized by NIST through the Mutual Recognition Agreement (CIRM MRA).
CALGAZ calibration devices were found to meet all applicable requirements of the National Highway Traffic Safety Administration Model Specifications for calibrating units for breath alcohol testers.

Manufactured Date: May 26, 2023

APPROVED BY: _____



We certify that all the cylinders for the Lot numbers identified herein are manufactured and tested within the requirements of CFR 49 part 178.65 and that physical and chemical test reports are on file and copies will be furnished upon request.

CALGAZ, a division of Airgas USA LLC
821 Chesapeake Drive, Cambridge, MD 21613-0149
Phone: (410) 228-6400 Fax: (410) 228-4251

DEPARTMENT OF
Traffic and Public Safety
This is to certify that

James G. Fillimon
New Jersey State Police

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 142 OF
THE LAWS OF 1966 IN THE OPERATION OF THE **ALCOTEST 9510**
A METHOD TO DETERMINE INTOXICATION
GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 3rd DAY OF October
TWO THOUSAND AND Twenty Three

[Signature]
COLONEL
NEW JERSEY STATE POLICE

[Signature]
ATTORNEY GENERAL
STATE OF NEW JERSEY

DEPARTMENT OF
Traffic and Public Safety
This is to certify that

James G. Fillimon
Breath Test Coordinator/Instructor

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 142 OF
THE LAWS OF 1966 IN THE OPERATION OF THE **Alcotest 9510**
A METHOD TO DETERMINE INTOXICATION
GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 18th DAY OF November
TWO THOUSAND AND Twenty Five

[Signature]
COLONEL
NEW JERSEY STATE POLICE

[Signature]
ATTORNEY GENERAL
STATE OF NEW JERSEY

ORIGINAL COURSE DATES

	DATE	Refresher Course PLACE	INSTRUCTOR
1.	4-10-25	MCFP	A.T.D.
2.			
3.			
4.			
5.			
6.			
7.			
8.			
9.			

S.P. 2938 (Rev. 10/22)

ORIGINAL COURSE DATES

	DATE	Refresher Course PLACE	INSTRUCTOR
1.			
2.			
3.			
4.			
5.			
6.			
7.			
8.			
9.			

S.P. 2938 (Rev. 10/22)

DEPARTMENT OF
Traffic and Public Safety
What is to certify that

Anthony P. Rubbe

Breath Test Coordinator/Instructor

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSIS PURSUANT TO CHAPTER 142 OF

THE LAWS OF 1963 IN THE OPERATION OF THE Alcotest 9510

A METHOD TO DETERMINE INTOXICATION

GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 22nd DAY OF January

TWO THOUSAND AND Twenty Five

[Signature]
COLONEL
NEW JERSEY STATE POLICE

[Signature]
ATTORNEY GENERAL
STATE OF NEW JERSEY

DEPARTMENT OF
Traffic and Public Safety
What is to certify that

Anthony P. Rubbe

New Jersey State Police

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSIS PURSUANT TO CHAPTER 142 OF

THE LAWS OF 1963 IN THE OPERATION OF THE Alcotest 9510

A METHOD TO DETERMINE INTOXICATION

GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 12th DAY OF October

TWO THOUSAND AND Twenty Three

[Signature]
COLONEL
NEW JERSEY STATE POLICE

[Signature]
ATTORNEY GENERAL
STATE OF NEW JERSEY

ORIGINAL COURSE DATES

	DATE	Refresher Course PLACE	INSTRUCTOR
1.			
2.			
3.			
4.			
5.			
6.			
7.			
8.			
9.			

S.P. 2938 (Rev. 01/13)

ORIGINAL COURSE DATES

	DATE	Refresher Course PLACE	INSTRUCTOR
1.	3-27-05	MCFA	<i>[Signature]</i>
2.			
3.			
4.			
5.			
6.			
7.			
8.			
9.			

S.P. 2938 (Rev. 10/22)

DEPARTMENT OF
Water and Public Safety
This is to certify that

Robert W. Waldrop

Breath Test Coordinator/Instructor

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 142 OF

THE LAWS OF 1946 IN THE OPERATION OF THE Alcotest 9510

A METHOD TO DETERMINE INTOXICATION.

GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 25th DAY OF March

TWO THOUSAND AND Twenty Four

[Signature]
 COLONEL
 NEW JERSEY STATE POLICE

[Signature]
 ATTORNEY GENERAL
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES _____

DATE	Refresher Course PLACE	INSTRUCTOR
1.		
2.		
3.		
4.		
5.		
6.		
7.		
8.		
9.		

S.P. 293B (Rev. 10/22)

DEPARTMENT OF
Water and Public Safety
This is to certify that

Robert W. Waldrop

New Jersey State Police

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 142 OF

THE LAWS OF 1946 IN THE OPERATION OF THE Alcotest 9510

A METHOD TO DETERMINE INTOXICATION.

GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 28th DAY OF April

TWO THOUSAND AND Twenty Three

[Signature]
 COLONEL
 NEW JERSEY STATE POLICE

[Signature]
 ATTORNEY GENERAL
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES _____

DATE	Refresher Course PLACE	INSTRUCTOR
1.		
2.		
3.		
4.		
5.		
6.		
7.		
8.		
9.		

S.P. 293B (Rev. 10/22)