

## ALCOTEST 9510 PARAMETER REPORT

### Equipment

Serial No.: ARMK-0097  
Firmware: 8326739 1.5  
WinCE application: 8326738 2.9  
Configuration: 8326737 3.10

Date: 10/13/2025  
Time: 06:54:40

### Parameter

min. blow time	5.0	s
min. breath volume for females of age 60+	1.2	L
min. breath volume for all other	1.5	L
min. blow flow	4.5	L/min
plateau detection limit	4	%
plateau detection start conc.	70	microgram/L
neg. flow detection (part. vacuum)	10.0	hPa
neg. flow detection sensitivity	10	
cal. gas abort volume	0.4	L
result-to-zero limit	0.0050	%BAC
ambient air check limit	0.0049	%BAC
interference det. d-criterion limit abs.	38	microgram/L
interference det. d-criterion limit rel.	10.0	%
interference det. t-criterion limit abs.	8	microgram/L
interference det. t-criterion limit rel.	2.1	%
IR CO2 offset	10	microgram/L
IR H2O offset	4	microgram/L
EC H2O offset	0	microgram/L
Value-based EC aging comp. on/off (1/0)	0	
Time-based EC aging comp. on/off (1/0)	1	
Time-based EC aging comp. per month	0.2	%
Time-based EC aging comp. maximum	3.0	%
EC fatigue comp. max. sum	15000	
EC fatigue comp. factor	50	
EC fatigue comp. minutes	180	
mouth alc. mark limit	500	
mouth alc. lower limit	30	
mouth alc. slope	6	
mouth alc. zero limit	50	
mouth alc. max. neg. sum	6	
mouth alc. max. 2nd derivative	35	

**ALCOTEST 9510 CERTIFICATION REPORT - WET ADJUST (PART I)**  
**South Brunswick**

**Equipment**

Inst. Model No.: ALCOTEST 9510 Serial No.: ARMK-0097  
Firmware: 8326739 1.5 Config.: 8326737 3.10 WinCE: 8326738 2.9

**Wet Adjust Record**

Wet Adjust File No.: 96 Wet Adjust Date: 10/13/2025 Wet Adjust No.: 6  
Wet Adjust Time: 07:39:39

Concentration: 0.100 %  
Adjusting Unit: X-Cal 2000 Adj. Unit Ser. No.: ARMA-0055 Adj. Unit Exp.: 08/19/2026  
Solution Lot No.: 24210 Soln. Bottle No.: 857 Adjust Soln. Exp.: 06/11/2026

Preadjust Simulator Temp.: 34.01 degree C  
Postadjust Simulator Temp.: 34.01 degree C

**Result**

**Procedure completed successfully.**

**Coordinator**

Last Name: RUBBE - First Name: ANTHONY MI: P Badge No.: 8352

On this date, I certified the above instrument in accordance with the Alcotest 9510 operator training and procedures established by the NJSP Office of Forensic Sciences.

*T.A. RUBBE 8352*

Signed:

Date: 10/13/2025

ID: 3018

**ALCOTEST 9510 CERTIFICATION REPORT - DRY ADJUST (PART II)**  
**South Brunswick**

**Equipment**

Inst. Model No.: ALCOTEST 9510 Serial No.: ARMK-0097  
Firmware: 8326739 1.5 Config.: 8326737 3.10 WinCE: 8326738 2.9

**Dry Adjust Record**

Dry Adjust File No.: 97 Dry Adjust Date: 10/13/2025 Dry Adjust No.: 6  
Dry Adjust Time: 08:03:10

Concentration: 0.100 %  
Dry Gas Lot No.: 302-402755160 Adjust Gas Exp.: 05/24/2026  
Barom. Model No.: Mensor CPG2300 Barom. Serial No.: 4100126W Barom. Cert. Exp.: 09/03/2026  
Pre-adjust Amb. Pressure: 1009 hPa Post-adjust Amb. Pressure: 1009 hPa

**Result**

**Procedure completed successfully.**

**Coordinator**

Last Name: RUBBE - First Name: ANTHONY MI: P Badge No.: 8352

On this date, I certified the above instrument in accordance with the Alcotest 9510 operator training and procedures established by the NJSP Office of Forensic Sciences.

*TR A. RUBBE*

Signed:

Date: 10/13/2025

ID: 3018

**ALCOTEST 9510 CERTIFICATION REPORT - LINEARITY (PART III)**  
**South Brunswick**

**Equipment**

Inst. Model No.: ALCOTEST 9510 Serial No.: ARMK-0097  
 Firmware: 8326739 1.5 Config.: 8326737 3.10 WinCE: 8326738 2.9

**Linearity Record**

Linearity File No.: 98 Lin. Date: 10/13/2025 Lin. No.: 6

0.040% Dry Gas Lot No.: 302-402730318 Adjust. Gas Exp.: 04/27/2026  
 0.080% Dry Gas Lot No.: 302-402732434 Adjust. Gas Exp.: 04/28/2026  
 0.160% Dry Gas Lot No.: 302-402922401 Adjust. Gas Exp.: 12/14/2026  
 0.300% Dry Gas Lot No.: 302-402757219 Adjust. Gas Exp.: 05/30/2026

**Data Summary**

Function	Result %BAC	Time hh:mm:ss	Barometric Pres. [hPa]	Comment(s) or Status Code
Ambient Air Blank	0.000	08:19:47		*TEST PASSED*
Control .04 Test 1 EC	0.039	08:20:22	1009	*TEST PASSED*
Control .04 Test 1 IR	0.039	08:20:22	1009	*TEST PASSED*
Ambient Air Blank	0.000	08:21:24		*TEST PASSED*
Control .04 Test 2 EC	0.040	08:21:36	1009	*TEST PASSED*
Control .04 Test 2 IR	0.040	08:21:36	1009	*TEST PASSED*
Ambient Air Blank	0.000	08:22:56		*TEST PASSED*
Control .08 Test 3 EC	0.078	08:23:29	1009	*TEST PASSED*
Control .08 Test 3 IR	0.079	08:23:29	1009	*TEST PASSED*
Ambient Air Blank	0.000	08:24:37		*TEST PASSED*
Control .08 Test 4 EC	0.080	08:24:49	1009	*TEST PASSED*
Control .08 Test 4 IR	0.080	08:24:49	1009	*TEST PASSED*
Ambient Air Blank	0.000	08:26:23		*TEST PASSED*
Control .16 Test 5 EC	0.156	08:26:56	1009	*TEST PASSED*
Control .16 Test 5 IR	0.158	08:26:56	1009	*TEST PASSED*
Ambient Air Blank	0.000	08:28:11		*TEST PASSED*
Control .16 Test 6 EC	0.160	08:28:23	1009	*TEST PASSED*
Control .16 Test 6 IR	0.159	08:28:23	1009	*TEST PASSED*
Ambient Air Blank	0.000	08:37:20		*TEST PASSED*
Control .30 Test 7 EC	0.303	08:37:58	1009	*TEST PASSED*
Control .30 Test 7 IR	0.303	08:37:58	1009	*TEST PASSED*
Ambient Air Blank	0.000	08:39:20		*TEST PASSED*
Control .30 Test 8 EC	0.309	08:39:35	1009	*TEST PASSED*
Control .30 Test 8 IR	0.307	08:39:35	1009	*TEST PASSED*
Ambient Air Blank	0.000	08:40:07		*TEST PASSED*

**Result**

**All tests within acceptable tolerance.**

**Coordinator**

Last Name: RUBBE - First Name: ANTHONY MI: P Badge No.: 8352

On this date, I certified the above instrument in accordance with the Alcotest 9510 operator training and procedures established by the NJSP Office of Forensic Sciences.

*TRM A.R.M. 8352*

Signed:

Date: 10/13/2025

ID: 3018

**ALCOTEST 9510 CYLINDER INSTALLATION REPORT - INLET 1**  
**South Brunswick**  
**SERIAL NUMBER: ARMK-0097**

**Equipment**

Inst. Model No.: ALCOTEST 9510 Serial No.: ARMK-0097  
 Firmware: 8326739 1.5 Config.: 8326737 3.10 WinCE: 8326738 2.9  
 Cyl1 Install File No.: 52 Cyl1 Install Date: 02/12/2025 Cyl1 Install No.: 2

**Control Tests (0.100%)**

Installation Inlet: #1 (Upper) Post test active Cyl.: #2 (Lower)  
 Dry Gas Lot No.: 302-403034219 Dry Gas Lot Exp.: 05/01/2027

**Data Summary**

Function	Result %BAC	Time hh:mm:ss	Barometric Pres. [hPa]	Comment(s) or Status Code
Ambient Air Blank	0.000	09:33:37		*TEST PASSED*
Control Test 1			1021	*TEST PASSED*
EC Result	0.097	09:34:23		*TEST PASSED*
IR Result	0.100	09:34:23		*TEST PASSED*
Ambient Air Blank	0.000	09:35:40		*TEST PASSED*
Control Test 2			1021	*TEST PASSED*
EC Result	0.100	09:36:04		*TEST PASSED*
IR Result	0.101	09:36:04		*TEST PASSED*
Ambient Air Blank	0.000	09:37:22		*TEST PASSED*
Control Test 3			1021	*TEST PASSED*
EC Result	0.101	09:37:46		*TEST PASSED*
IR Result	0.101	09:37:46		*TEST PASSED*
Ambient Air Blank	0.000	09:38:25		*TEST PASSED*

**Result**

All tests within acceptable tolerance.

**Coordinator**

Last Name: Waldrop - First Name: Robert MI: W Badge No.: 8256

On this date, I certified the above instrument in accordance with the Alcotest 9510 operator training and procedures established by the NJSP Office of Forensic Sciences.

*TR. Waldrop - 8256*

Signed:

Date: 02/12/2025

ID: 52

**ALCOTEST 9510 CYLINDER INSTALLATION REPORT - INLET 2**  
**South Brunswick**  
**SERIAL NUMBER: ARMK-0097**

**Equipment**

Inst. Model No.: ALCOTEST 9510 Serial No.: ARMK-0097  
 Firmware: 8326739 1.5 Config.: 8326737 3.10 WinCE: 8326738 2.9  
 Cyl2 Install File No.: 99 Cyl2 Install Date: 10/13/2025 Cyl2 Install No.: 2

**Control Tests (0.100%)**

Installation Inlet: #2 (Lower) Post test active Cyl.: #1 (Upper)  
 Dry Gas Lot No.: 302-403376678 Dry Gas Lot Exp.: 06/20/2028

**Data Summary**

Function	Result %BAC	Time hh:mm:ss	Barometric Pres. [hPa]	Comment(s) or Status Code
Ambient Air Blank	0.000	08:51:25		*TEST PASSED*
Control Test 1			1009	*TEST PASSED*
EC Result	0.099	08:52:11		*TEST PASSED*
IR Result	0.101	08:52:11		*TEST PASSED*
Ambient Air Blank	0.000	08:53:24		*TEST PASSED*
Control Test 2			1009	*TEST PASSED*
EC Result	0.101	08:53:48		*TEST PASSED*
IR Result	0.102	08:53:48		*TEST PASSED*
Ambient Air Blank	0.000	08:55:01		*TEST PASSED*
Control Test 3			1009	*TEST PASSED*
EC Result	0.102	08:55:25		*TEST PASSED*
IR Result	0.102	08:55:25		*TEST PASSED*
Ambient Air Blank	0.000	08:55:58		*TEST PASSED*

**Result**

**All tests within acceptable tolerance.**

**Coordinator**

Last Name: RUBBE - First Name: ANTHONY MI: P Badge No.: 8352

On this date, I certified the above instrument in accordance with the Alcotest 9510 operator training and procedures established by the NJSP Office of Forensic Sciences.

*TR. A. RUBBE 8352*

Signed:

Date: 10/13/2025

ID: 3018

# CERTIFICATE OF ANALYSIS

## EBS - ETHANOL BREATH STANDARD

Part Number: 4401036  
**DRAEGER MEDICAL SYSTEMS INC**

Sales order: 1130435101  
 Date: May 28, 2024

METHOD OF ANALYSIS: IR Breath Alcohol Analyzer  
 ANALYTICAL ACCURACY: +/-0.002 BrAC or +/-2% whichever is greater.  
 CALGAZ LOT#: 302-403034219  
 ETHANOL IN NITROGEN

Manufactured Date: May 01, 2024  
 Product Expiration: May 01, 2027

COMPONENT	PPM	( BrAC )
ETHANOL	260.5PPM	(0.100)
NITROGEN	BAL	
AVERAGE ANALYTICAL VALUE	PPM	( BrAC )
ETHANOL	263.0	(0.101)

REFERENCE STANDARD	CYLINDER	CONCENTRATION PPM
N.M.I. TRACEABLE STANDARDS*	ND38424	260.7

\* CERTIFICATION TRACEABLE TO NATIONAL METROLOGY INSTITUTE TRACEABLE STANDARDS

**TRACEABILITY**

**Preparation:**

Gas mixtures manufactured with balances calibrated by an ISO 17025 accredited company using NIST traceable weights and meets or exceeds the requirements of NIST Handbook 44.

Traceable certificate numbers 3445312 and 3398673.

**Analytical:**

Analytical Instruments Calibrated Using NMI Traceable Standards.


Certification Numbers: A679-20190918, D049803-20220329

No effecting environmental conditions during analysis.

\*NMI is recognized by NIST through the Mutual Recognition Agreement (CIPM MRA).

CALGAZ calibration devices were found to meet all applicable requirements of the National Highway Traffic Safety Administration Model Specifications for calibrating units for breath alcohol testers.

APPROVED BY: \_\_\_\_\_



"We certify that all the cylinders for the Lot numbers identified herein are manufactured and tested within the requirements of CFR 49 part 178.65 and that physical and chemical test reports are on file and copies will be furnished upon request."

**CALGAZ, a division of Airgas USA LLC**  
 821 Chesapeake Drive, Cambridge, MD 21613-0149  
 Phone: (410) 228-6400 Fax: (410) 228-4251

# CERTIFICATE OF ANALYSIS

## EBS - ETHANOL BREATH STANDARD

Part Number: 4401036  
DRAEGER MEDICAL SYSTEMS INC

Sales order: 1141350717  
Date: July 25, 2025

METHOD OF ANALYSIS: IR Breath Alcohol Analyzer  
ANALYTICAL ACCURACY: +/-0.002 BrAC or +/-2% whichever is greater.  
CALGAZ LOT#: 302-403376678  
ETHANOL IN NITROGEN

Manufactured Date: June 20, 2025  
Product Expiration: June 20, 2028

COMPONENT	PPM	( BrAC )
ETHANOL	260.5PPM	(0.100)
NITROGEN	BAL	
AVERAGE ANALYTICAL VALUE	PPM	( BrAC )
ETHANOL	262.6	(0.101)

REFERENCE STANDARD	CYLINDER	CONCENTRATION PPM
N.M.I. TRACEABLE STANDARDS*	ND49826	260.1

\* CERTIFICATION TRACEABLE TO NATIONAL METROLOGY INSTITUTE TRACEABLE STANDARDS

#### TRACEABILITY

##### Preparation:

Gas mixtures manufactured with balances calibrated by an ISO 17025 accredited company using NIST traceable weights and meets or exceeds the requirements of NIST Handbook 44.  
Traceable certificate numbers 3445312 and 3398673.


##### Analytical:

Analytical Instruments Calibrated Using NMI Traceable Standards.  
Certification Numbers: A679-20190918, D049803-20220329

No effecting environmental conditions during analysis.


\*NMI is recognized by NIST through the Mutual Recognition Agreement (CIPM MRA).

CALGAZ calibration devices were found to meet all applicable requirements of the National Highway Traffic Safety Administration Model Specifications for calibrating units for breath alcohol testers.

APPROVED BY: 

"We certify that all the cylinders for the Lot numbers identified herein are manufactured and tested within the requirements of CFR 49 part 178.65 and that physical and chemical test reports are on file and copies will be furnished upon request."

**CALGAZ, a division of Airgas USA LLC**  
821 Chesapeake Drive, Cambridge, MD 21613-0149  
Phone: (410) 228-6400 Fax: (410) 228-4251



**Dräger**

**Alcotest 9510**

**CERTIFICATE OF ACCURACY**

This is to certify that the Alcotest 9510 has been tested for accuracy and found to be in compliance with the National Highway Traffic Safety Administration Standard for evidential breath testing devices. The Alcotest MKIII-C is compliant as a "mobile" and "nonmobile" EBT with 49 FR 48854, 49 FR 48864, and 58 FR 48705. The manufacturer recommends accuracy verification of this instrument within 12 months of the calibration date below, or sooner, according to your state's specifications.

Certification Date:

Serial Number:

11/28/2023

ARMK-0097

Draeger, Inc.

JP

MB



State of New Jersey

OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
POST OFFICE BOX 7068
WEST TRENTON, NJ 08628-0068
(609) 882-2000

PHILIP D. MURPHY
Governor

TAHESHA L. WAY
Lt. Governor

MATTHEW J. PLATKIN
Attorney General

COLONEL PATRICK J. CALLAHAN
Superintendent

CERTIFICATION OF ANALYSIS
0.100 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.1174 to 0.1246 grams per 100 milliliters of solution.

MANUFACTURER: Draeger, Inc.

ANALYSIS DATE: 07/18/2024

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 24210

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1195 to 0.1217 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is June 11, 2026.

As OFS Director for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Michael Kennedy
Michael Kennedy
Director
NJSP Office of Forensic Sciences

Sworn to and subscribed before me this 24 day of July, 2024.

Notary

KAREN E. STAHL
NOTARY PUBLIC OF NEW JERSEY
Commission # 50110522
My Commission Expires 6/13/2024

KAREN E. STAHL
NOTARY PUBLIC OF NEW JERSEY
Commission # 50110522

KAREN E. STAHL
NOTARY PUBLIC OF NEW JERSEY
Commission # 50110522



An Internationally Accredited Agency

New Jersey Is An Equal Opportunity Employer
Printed on Recycled Paper and Recyclable



Customer: DRAEGER INC  
7256 S SAM HOUSTON PKWY W  
STE 100  
HOUSTON, TX 77085  
PO Number: SUC4303700862



Certificate/SO Number: 5-F8B2G-400-1 Revision 0

Manufacturer: Draeger Safety AG & Co. KGaA  
Model Number: X-Cal 2000  
Description: Breath Alcohol Simulator  
Serial Number: ARMA-0055  
ID: NONE

As-Found: In Tolerance  
As-Left: In Tolerance  
Issue Date: Aug 19, 2025  
Calibration Date: Aug 19, 2025  
Due Date: Aug 19, 2026

Calibrated To: Manufacturer Specification  
Calibration Procedure: 1-AC103519-2

Transcal Calibration Laboratories have been audited and found in compliance with ISO/IEC 17025:2017. Accredited calibrations performed within the Lab Scope of Accreditation are indicated by the presence of the Accrediting Body Logo and Certificate Number. Any measurements on an accredited calibration not covered by the Lab Scope of Accreditation are listed in the notes section of the certificate. SCC, NRC, CLAS or ANAB do not guarantee the accuracy of an individual calibration by accredited laboratories.

Transcal calibrations, as applicable, are performed in compliance with the requirements of the Transcal Quality Manual QAC-P01-000, the customer Purchase Order and/or Quality Agreement requirements, ISO 9001:2015, ANSI/NCSL 2540.1-1994 (R2002), and ISO 10012:2003, as applicable. When specified contractually, the requirements of ISO TS 16949:2009, IOCFR21, IOCFR50 App. B, ASME NQA-1:2012, and ANSI/NCSL 2540.3:2005 (R2013) are also covered.

Complete records of work performed are maintained by Transcal and are available for inspection. Laboratory standards used in the performance of this calibration are listed on this certificate.


Transcal documents the traceability of measurements to the SI units through the National Institute of Standards and Technology (NIST), or the National Research Council of Canada (NRC), or other national measurement institutes (NMI) that are signatories to the CIPM Mutual Recognition Arrangement, or accepted fundamental and/or natural physical constants, or by the use of specified methods, consensus standards or ratio type measurements. Documentation supporting traceability information is available for review upon written request at a Transcal facility. The measured quantity and the measurement uncertainty are required for further dissemination of traceability.

Uncertainties are reported with a coverage factor k=2, providing a level of confidence of approximately 95%. All calibrations have been performed using processes having a TUR of 4:1 or better (5:1 for mass calibrations), unless otherwise noted. The Test Uncertainty Ratio (TUR) is calculated in accordance with NCCL International RP-14. For mass calibrations, conventional mass reference masses referred to B.D. gram.

The results in this report relate only to the item calibrated or tested. Recorded calibration data is valid at the time of calibration within the stated uncertainties at the environmental conditions noted. The determination of compliance to the specification is specific to the model/serial no./ID no. referenced above based on the tolerances shown; these tolerances are either the original equipment manufacturer's (OEM's) warranted specifications or the client's requested specifications. Any number of factors can cause a unit to drift out of tolerance at any time following its calibration. Limitations on the uses of this instrument are detailed in the OEM's operating instructions. This certificate may not be reproduced except in full, without the written approval of Transcal. Additional information, if applicable may be included on separate report(s).

# CALIBRATED BY TRANSOT

# CERTIFICATE OF CALIBRATION



Customer: DRAEGER INC  
 7256 S SAM HOUSTON PKWY W  
 STE 100  
 HOUSTON, TX 77085  
 P.O. Number: SUC4303700862

Certificate/SO Number: 5-F8B2G-400-1 Revision 0

### As Found/As Left Data

Description	Setpoints	Accuracy	As Found / As Left		High Limit	Low Limit	O	Cal Process		Units	TUR
			O	T				Uncertainty (k=2, 3)	Measurement Uncertainty (k=2, 3)		
<b>Function Checks</b>											
Bubble Check					P						
Seal Check					P						
Temperature Source: Accuracy Test											
Accuracy Test	34.00°C	±(0.02 °C)		34.02	34.00 °C			1.5e-002	1.6e-002	°C	1.3:1
Temperature Source: Stability Test											
Stability Test	0.00°C	±(0.02 °C)		0.02	0.00 °C			1.5e-002	1.6e-002	°C	1.3:1

Field not applicable.

### Traceable Standards

Asset	Manufacturer	Model Number	Description	Cal Date	Due Date	Traceability Number	Use
05H1479	AccuMac Corporation	AM1760-12-S	Secondary SFRT	12-Aug-24	31-Aug-25	H32XM-4-1	AF/AL
HP927312	Hart Scientific/Fluke	1575	Super Thermometer	10-Jul-24	31-Jan-26	5-3-HP927312-9-1	AF/AL

The use of the standard is defined as: AF - used for as-found readings, AL - used for as-left readings.

### Environmental Data


Temperature	Relative Humidity	Temp / RH Asset	Lab Area	Lab Description
68.84°F / 20.52°C	55.99%	DswK11	G	Temperature

### Decision Rule

When compliance statements are present, they are reported without indicating in the effects of uncertainty and comply with the guidelines as follows: The acceptance zone is defined as: less than or equal to the high limit, and/or: greater than or equal to the low limit. The rejection zones are defined as: greater than the high limit and/or less than the low limit. Single measurement results in the acceptance zone

**CALIBRATED**  
BY **TRANSOCT**


# CERTIFICATE OF CALIBRATION



Customer: DRAEGER INC  
7256 S SAM HOUSTON PKWY W  
STE 100  
HOUSTON, TX 77085  
PO Number: SUC4303700862

Certificate/SO Number: 5-F8B2G-400-1 Revision 0

are identified as in-tolerance. Single measurement results in the rejection zone are identified as out-of-tolerance (OOT). When all measurement results are in the acceptance zone for repeated measurements, for the same characteristic, the test is identified as in-tolerance. For repeated characteristic measurements, a single measurement result in the rejection zone, will cause the test to be identified as out-of-tolerance (OOT). Data rejection for cause, (oullars) is permitted after the "Determining and Verifying Out Of Tolerance (OOT) and/or Op Fail Readings" procedure outlined in this document has been completed and the anomalous reading cannot be repeated, and the anomalous reading does not represent the system under test. Statements of conformity are binary.



Customer: DRAEGER INC  
7266 S SAM HOUSTON PKWY W  
STE 100  
HOUSTON, TX 77085  
PO Number: SUC4303700862


Certificate/SO Number: 5-F8B2G-400-1 Revision 0

**Legend**

Topic	Description
Accuracy	UUT specification that establishes expected tolerances and a time limit (calibration interval) over which the instrument is expected to hold these tolerances
As Found	Initial measurement results
As Left	Measurement results after adjustment and/or repair
Blank Data Field	Test is not applicable for the UUT
Cal Process Uncertainty (CPU)	The uncertainty of calibration process for the reported measurement result
Calibration Date	Indicates the date that the calibration was completed
Cover Factor (k)	A measure of uncertainty that defines an interval about the measurement result
Due Date	Indicates the end of the calibration cycle as requested by the customer
Issue Date	Indicates the date that the calibration has passed the Data Review Process and was signed by an authorized signatory or the date that a revision to the original certificate has been issued
Low / High Limits	Establishes UUT acceptable performance limits for the test measurement
Measurement Uncertainty	The dispersion of the values attributed to a measured quantity
OGA	Out of Acceptance (H)
OOT	Out of Tolerance (T)
Setpoints	Measurement target values
Traceability	Unbroken chain of comparisons relating an instrument's measurements to a known standard(s)
Traceability Number	Unique identifier(s) used to document traceability of calibration standards
TUR	Test Uncertainty Ratio, ratio of the tolerance or specification of the test measurement in relation to the uncertainty in measurement results
UUT	Unit Under Test

**CALIBRATED**  
BY TRANSCAT

# CERTIFICATE OF CALIBRATION



Customer: DRAEGER INC  
7256 S SAM HOUSTON PKWY W  
STE 100  
HOUSTON, TX 77085  
P.O Number: SUC4303700862

Certificate/SO Number: 5-F8B2G-400-1 Revision 0

Calibrated At:  
18115 Park Row  
Houston, TX 77084

Facility Responsible:  
18115 Park Row  
Houston, TX 77084  
800-628-1470

Unit Barcode:  09005541600

Date Received: August 06, 2025  
Service Level: RD

Calibrated By:  
 Joe Martinez  
Joe Martinez  
Calibration Technician


Electronically Signed By:  
Aug 16, 2025  
09:05:04 -04:00

Reviewed By:  
 Josh Solaire  
Josh Solaire  
Lab Manager

Electronically Signed By:  
Aug 10, 2025  
14:15:35 -04:00  
Cynthia Walker for

Certificate - Page 5 of 5  
Reprinted on August 26, 2025

Customer Number: 1-659111-000  
OPS-F20-Q14R11 07/27/23 FP-00139-09/2021



Customer: DRAEGER INC  
7256 S SAM HOUSTON PKWY W  
STE 100  
HOUSTON, TX 77035  
PO Number: SUC4303700862

Certificate/SO Number: 5-F8B2G-160-1 Revision 0

Manufacturer: Wika Instr/Mensor Corp/Trend  
Model Number: CP32300  
Description: Portable Barometer  
Serial Number: 4100126W  
ID: NONE

As-Found: Out Of Tolerance  
As-Left: In Tolerance

Issue Date: Sep 03, 2025  
Calibration Date: Sep 03, 2025  
Due Date: Sep 03, 2026

Calibrated To: Manufacturer Specification

Calibration Procedure: 1-AC107288-0

Transcat Calibration Laboratories have been audited and found in compliance with ISO/IEC 17025:2017. Accredited calibrations performed within this Lab's Scope of Accreditation are indicated by the presence of the Accrediting Body Logo and Certificate Number. Any measurements on an accredited calibration not covered by the Lab's Scope of Accreditation are listed in the notes section of the certificate. SCC, NRC, CLAS or ANAB do not guarantee the accuracy of an individual calibration by accredited laboratories.

Transcat calibrations, as applicable, are performed in compliance with the requirements of the Transcat Quality Manual QAC-P01-000, the customer Purchase Order and/or Quality Agreement requirements, ISO 9001:2015, ANSI/NCSL Z540.1-1994 (R2002), and ISO 10012:2003, as applicable, when specified contractually, the requirements of ISO TS 16949:2009, IOCFR21, IOCFR60 App. B, ASME NQA-1:2012, and ANSI/NCSL Z540.3-2006 (R2013) are also covered.

Complete records of work performed are maintained by Transcat and are available for inspection. Laboratory standards used in the performance of this calibration are listed on this certificate.

Transcat documents the traceability of measurements to the SI units through the National Institute of Standards and Technology (NIST), or the National Research Council of Canada (NRC), or other national measurement institutes (NMI) that are signatories to the CIPM Mutual Recognition Arrangement, or accepted fundamental and/or natural physical constants, or by the use of specified methods, consensus standards or ratio type measurements. Documentation supporting traceability information is available for review upon written request at a Transcat facility. The measured quantity and the measurement uncertainty are required for further dissemination of traceability.


Uncertainties are reported with a coverage factor k=2, providing a level of confidence of approximately 95%. All calibrations have been performed using processes having a TUR of 4:1 or better (3:1 for mass calibrations), unless otherwise noted. The Test Uncertainty Ratio (TUR) is calculated in accordance with NCSL International RP-10. For mass calibrations: Conventional mass referenced to 9.0 gram.

The results in this report relate only to the item calibrated or tested. Recorded calibration data is valid at the time of calibration within the stated uncertainties at the environmental conditions noted. The determination of compliance to the specification is specific to the model/serial no./ID no. referenced above based on the tolerances shown; these tolerances are either the original equipment manufacturer's (OEM's) warranted specifications of the client's requested specifications. Any number of factors can cause a unit to drift out of tolerance at any time following its calibration. Limitations on the uses of this instrument are detailed in the OEM's operating instructions. This certificate may not be reproduced except in full, without the written approval of Transcat. Additional information, if applicable may be included on separate report(s).

Notes:

Unit received out of tolerance. Adjustments were made to meet customer specs. The OOT readings were verified.

Customer: DRAEGER INC  
 7266 S SAM HOUSTON PKWY W  
 STE 100  
 HOUSTON, TX 77065  
 PO Number: SUC4303700662




Certificate/SO Number: 5-F8B2G-160-1 Revision 0

As Found Data

Description	Seipoints	Accuracy	As Found		Cal Process Uncertainty (k=2; E)	Measurement Uncertainty (k=2; E)	Units	TUR
			Low Limit	High Limit				
Pressure Measure: 552 to 1172 mbara Range								
	550.10mbara	±(0.015% FS)	549.82	550.28	1.0e-002	1.2e-002	mbara	17.2 : 1
	610.03mbara	±(0.015% FS)	609.85	610.21	1.2e-002	1.3e-002	mbara	15.5 : 1
	680.39mbara	±(0.015% FS)	680.21	680.57	1.3e-002	1.4e-002	mbara	13.9 : 1
	734.30mbara	±(0.015% FS)	734.12	734.48	1.4e-002	1.5e-002	mbara	12.9 : 1
	804.66mbara	±(0.015% FS)	804.48	804.84	1.5e-002	1.6e-002	mbara	11.8 : 1
	864.93mbara	±(0.015% FS)	864.75	865.11	1.6e-002	1.7e-002	mbara	11.0 : 1
	924.93mbara	±(0.015% FS)	924.75	925.11	1.8e-002	1.8e-002	mbara	10.2 : 1
	985.23mbara	±(0.015% FS)	985.05	985.41	1.9e-002	2.0e-002	mbara	9.6 : 1
	1043.9mbara	±(0.015% FS)	1043.7	1044.1	2.0e-002	2.1e-002	mbara	10.1 : 1
	1114.2mbara	±(0.015% FS)	1114.0	1114.4	2.1e-002	2.2e-002	mbara	9.4 : 1
	1174.6mbara	±(0.015% FS)	1174.4	1174.8	2.2e-002	2.3e-002	mbara	9.0 : 1
	924.93mbara	±(0.015% FS)	924.75	925.11	1.8e-002	1.8e-002	mbara	10.2 : 1
	864.93mbara	±(0.015% FS)	864.75	865.11	1.8e-002	1.7e-002	mbara	11.0 : 1
	804.65mbara	±(0.015% FS)	804.47	804.83	1.5e-002	1.6e-002	mbara	11.8 : 1

Customer: DRAEGER INC  
 7256 S SAM HOUSTON PKWY W  
 STE 100  
 HOUSTON, TX 77065  
 PO Number: SUC4303700862



## Certificate/SO Number: 5-F8B2G-160-1 Revision 0

### As Left Data

Description	Setpoints	Accuracy	Low Limit	High Limit	As Left	Cal Process Uncertainty (k=2; s)	Measurement Uncertainty (k=2; s)	Units	TUR
Pressure Measure: 582 to 1172 mbara Range									
	550.00mbara	±(0.015% FS)	549.90	550.26	550.00 mbara	1.0e-002	1.2e-002	mbara	17.2 : 1
	610.02mbara	±(0.015% FS)	609.84	610.20	609.90 mbara	1.2e-002	1.3e-002	mbara	16.5 : 1
	680.38mbara	±(0.015% FS)	680.20	680.56	680.30 mbara	1.3e-002	1.4e-002	mbara	13.9 : 1
	734.28mbara	±(0.015% FS)	734.10	734.46	734.20 mbara	1.4e-002	1.5e-002	mbara	12.9 : 1
	804.64mbara	±(0.015% FS)	804.46	804.82	804.60 mbara	1.5e-002	1.6e-002	mbara	11.8 : 1
	864.92mbara	±(0.015% FS)	864.74	865.10	864.90 mbara	1.6e-002	1.7e-002	mbara	11.0 : 1
	924.92mbara	±(0.015% FS)	924.74	925.10	924.90 mbara	1.8e-002	1.8e-002	mbara	10.2 : 1
	985.22mbara	±(0.015% FS)	985.04	985.40	985.30 mbara	1.9e-002	2.0e-002	mbara	9.6 : 1
	1043.6mbara	±(0.015% FS)	1043.6	1044.0	1043.9 mbara	2.0e-002	2.1e-002	mbara	10.1 : 1
	1114.2mbara	±(0.015% FS)	1114.0	1114.4	1114.3 mbara	2.1e-002	2.2e-002	mbara	9.4 : 1
	1174.6mbara	±(0.015% FS)	1174.4	1174.8	1174.6 mbara	2.2e-002	2.3e-002	mbara	9.0 : 1
	924.92mbara	±(0.015% FS)	924.74	925.10	924.90 mbara	1.8e-002	1.8e-002	mbara	10.2 : 1
	864.92mbara	±(0.015% FS)	864.74	865.10	864.90 mbara	1.6e-002	1.7e-002	mbara	11.0 : 1
	804.64mbara	±(0.015% FS)	804.46	804.82	804.60 mbara	1.5e-002	1.6e-002	mbara	11.8 : 1


Field not applicable.

### Traceable Standards

Asset	Manufacturer	Model Number	Description	Cal Date	Due Date	Traceability Number	Use
DW11BA	Fluke/DH Instruments	PG7601	Piston Gauges	31-Jul-25	31-Jul-26	5-8DW11BA-20-1	AF/AL
DW11CA	Fluke/DH Instruments	MS-AMH-38	AMH Mass Set	5-Jun-25	30-Sep-25	5-8DW11CA-40-1	AF/AL
DW11LOW	Fluke/DH Instruments	PC-71007/600-10-TC	Gas Piston-Cylinder Module	8-Apr-22	30-Apr-27	5-8DW11LOW-3-1	AF/AL
DW11MASS	Fluke/DH Instruments	MS-AMH-38	AMH Mass Set	5-Mar-25	31-Mar-26	5-8DW11MASS-12-1	AF/AL

The use of the standard is defined as: AF - used for as-found readings, AL - used for as-left readings.

Customer: DRAEGER INC  
 7256 S SAM HOUSTON PKWY W  
 STE 100  
 HOUSTON, TX 77085  
 PO Number: SUC4303700062



Certificate/SO Number: 5-F8B2G-160-1 Revision 0


### Environmental Data

Temperature	Relative Humidity	Temp / RH Asset	Lab Area	Lab Description
72.72°F / 22.62°C	44.00%	DewK9	B	GP Pressure

### Decision Rule

When compliance statements are present, they are reported without factoring in the effects of uncertainty and comply with the guidelines as follows: The acceptance zone is defined as: less than or equal to the high limit, and/or greater than or equal to the low limit. The rejection zones are defined as greater than the high limit and/or less than the low limit. Single measurement results in the acceptance zone are identified as in-tolerance. Single measurement results in the rejection zone are identified as out-of-tolerance (OOT). When all measurement results are in the acceptance zone for repeated measurements, for the same characteristic, the test is identified as in-tolerance. For repeated characteristic measurements, a single measurement result in the rejection zone, will cause the test to be identified as out-of-tolerance (OOT). Data rejection for cause, (outliers) is permitted after the "Determining and Verifying Out Of Tolerance (OOT) and/or Op Fail Readings" procedure outlined in this document has been completed and the anomalous reading cannot be repeated, and the anomalous reading does not represent the system under test. Statements of conformity are binary.

Customer: DRAEGER INC  
7256 S SAM HOUSTON PIKWY W  
SITE 100  
HOUSTON, TX 77085  
PO Number: SUC/303700862



Certificate/SO Number: 5-F8B2G-160-1 Revision 0


**Legend**

Topic	Description
Accuracy	UUT specification that establishes expected tolerances and a time limit (calibration interval) over which the instrument is expected to hold these tolerances
As Found	Initial measurement results
As Left	Measurement results after adjustment and/or repair
Blank Data Field	Test is not applicable for the UUT
Cal Process Uncertainty (CPU)	The uncertainty of calibration process for the reported measurement result
Calibration Date	Indicates the date that the calibration was completed
Cover Factor (K)	A measure of uncertainty that defines an interval about the measurement result
Due Date	Indicates the end of the calibration cycle as requested by the customer
Issue Date	Indicates the date that the calibration has passed the Data Review Process and was signed by an authorized signatory or the date that a revision to the original certificate has been issued
Low / High Limits	Establishes URIT acceptable performance limits for the test measurement
Measurement Uncertainty	The dispersion of the values attributed to a measured quantity
COA	Out of Acceptance (f)
OOT	Out of Tolerance (°)
Setpoints	Measurement target values
Traceability	Unbroken chain of comparisons relating an instrument's measurements to a known standard(s)
Traceability Number	Unique Identifier(s) used to document traceability of calibration standards
TUR	Test Uncertainty Ratio, ratio of the tolerance or specification of the test measurement in relation to the uncertainty in measurement results
UUT	Unit Under test

**CALIBRATED**  
BY TRANSCOT

# CERTIFICATE OF CALIBRATION

Customer: DRAEGER INC  
7256 S SAM HOUSTON PKWY W  
STE 100  
HOUSTON, TX 77065  
PO Number: SUC4303700862



Certificate/SO Number: 5-F8B2G-160-1 Revision 0

Calibrated At:  
16115 Park Row  
Houston, TX 77064

Facility Responsible:  
16115 Park Row  
Houston, TX 77064  
800-828-1470

Unit Barcode:  
09005541233

Date Received: August 05, 2025  
Service Level: R0

Calibrated By:  
  
Alex Spilker  
Calibration Technician  
Sep 03, 2025 01:30:28 -0400

Reviewed By:  
  
Graham Walker  
Lab Manager  
Sep 05, 2025 05:53:22 -0400

Electronically Signed By:  
Graham Walker for

Certificate - Page 6 of 6  
Reprinted on September 04, 2025

Customer Number: 1-65911-000  
OPS-F20-019R11 07/27/23 FP00-IR9 10/2/21

# CERTIFICATE OF ANALYSIS

## EBS - ETHANOL BREATH STANDARD

DEPT OF LAW AND PUBLIC SAETY

Sales order: 1120654933

Date: May 30, 2023

METHOD OF ANALYSIS: IR Breath Alcohol Analyzer  
ANALYTICAL ACCURACY: +/-0.002 BrAC or +/-2% whichever is greater.  
CALGAZ LOT#: 302-402755160  
ETHANOL IN NITROGEN

Product Expiration: May 24, 2026

COMPONENT	PPM	( BrAC )
ETHANOL	260.5PPM	(0.100)
NITROGEN	BAL	
AVERAGE ANALYTICAL VALUE	PPM	( BrAC )
ETHANOL	261.6	(0.100)

REFERENCE STANDARD	CYLINDER	CONCENTRATION PPM
N.M.I. TRACEABLE STANDARDS*	ND38424	260.7

\* CERTIFICATION TRACEABLE TO NATIONAL METROLOGY INSTITUTE TRACEABLE STANDARDS.

### TRACEABILITY

#### Preparation:

Gas mixtures manufactured with balances calibrated by an ISO 17025 accredited company using NIST traceable weights and meets or exceeds the requirements of NIST Handbook 44.

Traceable certificate numbers 3445312 and 3398673.

#### Analytical:

Analytical Instruments Calibrated Using NMI Traceable Standards.


Certification Numbers: A679-20190918, D049803-20220329

No effecting environmental conditions during analysis.

\*NMI is recognized by NIST through the Mutual Recognition Agreement (CIPM MRA).

CALGAZ calibration devices were found to meet all applicable requirements of the National Highway Traffic Safety Administration Model Specifications for calibrating units for breath alcohol testers.

Manufactured Date: May 24, 2023

APPROVED BY: 

"We certify that all the cylinders for the Lot numbers identified herein are manufactured and tested within the requirements of CFR 49 part 178.65 and that physical and chemical test reports are on file and copies will be furnished upon request."

**CALGAZ, a division of Airgas USA LLC**

821 Chesapeake Drive, Cambridge, MD 21613-0149

Phone: (410) 228-6400

Fax: (410) 228-4251

# CERTIFICATE OF ANALYSIS

## EBS - ETHANOL BREATH STANDARD

DEPT OF LAW AND PUBLIC SAFETY

Sales order: 1120656707

Date: May 30, 2023

METHOD OF ANALYSIS: IR Breath Alcohol Analyzer

ANALYTICAL ACCURACY: +/-0.002 BrAC or +/-2% whichever is greater:

CALGAZ LOT#: 302-402730318

ETHANOL IN NITROGEN

Product Expiration: April 27, 2026

COMPONENT	PPM	(BrAC)
ETHANOL	104.2PPM	(0.040)
NITROGEN	BAL	
AVERAGE ANALYTICAL VALUE	PPM	(BrAC)
ETHANOL	107.8	(0.041)

REFERENCE STANDARD	CYLINDER	CONCENTRATION PPM
N.M.I. TRACEABLE STANDARDS*	ND38424	260.7

\* CERTIFICATION TRACEABLE TO NATIONAL METROLOGY INSTITUTE TRACEABLE STANDARDS:

### TRACEABILITY

#### Preparation:

Gas mixtures manufactured with balances calibrated by an ISO 17025 accredited company using NIST traceable weights and meets or exceeds the requirements of NIST Handbook 44.

Traceable certificate numbers 3445312 and 3398673.

#### Analytical:

Analytical Instruments Calibrated Using NMI Traceable Standards.


Certification Numbers: A679-20190918, D049803-20220329

No effecting environmental conditions during analysis.

\*NMI is recognized by NIST through the Mutual Recognition Agreement (CIPM MRA).

CALGAZ calibration devices were found to meet all applicable requirements of the National Highway Traffic Safety Administration Model Specifications for calibrating units for breath alcohol testers.

Manufactured Date: April 27, 2023

APPROVED BY: 

\*We certify that all the cylinders for the Lot numbers identified herein are manufactured and tested within the requirements of CFR 49 part 178.65 and that physical and chemical test reports are on file and copies will be furnished upon request.\*

**CALGAZ, a division of Airgas USA LLC**  
821 Chesapeake Drive, Cambridge, MD 21613-0149  
Phone: (410) 228-6400 Fax: (410) 228-4251

# CERTIFICATE OF ANALYSIS

## EBS - ETHANOL BREATH STANDARD

Sales order: 1120656618  
Date: May 25, 2023

DEPT OF LAW AND PUBLIC SAFETY

METHOD OF ANALYSIS: IR Breath Alcohol Analyzer  
 ANALYTICAL ACCURACY: +/-0.002 BrAC or +/-2% whichever is greater.  
 CALGAZ LOT#: 302-402732434  
 ETHANOL IN NITROGEN

Product Expiration: April 28, 2026

COMPONENT	PPM	( BrAC )
ETHANOL	208.4PPM	(0.080)
NITROGEN	BAL	
<hr/>		
AVERAGE ANALYTICAL VALUE	PPM	( BrAC )
ETHANOL	210.4	(0.081)

REFERENCE STANDARD	CYLINDER	CONCENTRATION PPM
N.M.I. TRACEABLE STANDARDS*	ND38424	260.7

\* CERTIFICATION TRACEABLE TO NATIONAL METROLOGY INSTITUTE TRACEABLE STANDARDS

**TRACEABILITY**

**Preparation:**

Gas mixtures manufactured with balances calibrated by an ISO 17025 accredited company using NIST traceable weights and meets or exceeds the requirements of NIST Handbook 44.

Traceable certificate numbers 3445312 and 3398673.

**Analytical:**

Analytical Instruments Calibrated Using NMI Traceable Standards.  
 Certification Numbers: A679-20190918, D049803-20220329

No effecting environmental conditions during analysis.

\*NMI is recognized by NIST through the Mutual Recognition Agreement (CIPM MRA).  
 CALGAZ calibration devices were found to meet all applicable requirements of the National Highway Traffic Safety Administration Model Specifications for calibrating units for breath alcohol testers.

Manufactured Date: April 28, 2023

APPROVED BY: \_\_\_\_\_



"We certify that all the cylinders for the Lot numbers identified herein are manufactured and tested within the requirements of CFR 49 part 178.65 and that physical and chemical test reports are on file and copies will be furnished upon request."

**CALGAZ, a division of Airgas USA LLC**  
 821 Chesapeake Drive, Cambridge, MD 21613-0149  
 Phone: (410) 228-6400 Fax: (410) 228-4251

# CERTIFICATE OF ANALYSIS

## EBS - ETHANOL BREATH STANDARD

Part Number: 4401040NJ  
DRAEGER MEDICAL SYSTEMS INC

Sales order: 1126209454  
Date: December 18, 2023

METHOD OF ANALYSIS: IR Breath Alcohol Analyzer  
ANALYTICAL ACCURACY: +/-0.002 BrAC or +/-2% whichever is greater.  
CALGAZ LOT#: 302-402922401  
ETHANOL IN NITROGEN

Product Expiration: December 14, 2026

COMPONENT	PPM	( BrAC )
ETHANOL	416.8PPM	(0.160)
NITROGEN	BAL	
AVERAGE ANALYTICAL VALUE	PPM	( BrAC )
ETHANOL	418.6	(0.161)

REFERENCE STANDARD	CYLINDER	CONCENTRATION PPM
N.M.I. TRACEABLE STANDARDS*	ND38424	260.7

\* CERTIFICATION TRACEABLE TO NATIONAL METROLOGY INSTITUTE TRACEABLE STANDARDS

### TRACEABILITY

#### Preparation:

Gas mixtures manufactured with balances calibrated by an ISO 17025 accredited company using NIST traceable weights and meets or exceeds the requirements of NIST Handbook 44.

Traceable certificate numbers 3445312 and 3398673.

#### Analytical:

Analytical Instruments Calibrated Using NMI Traceable Standards.


Certification Numbers: A679-20190918, D049803-20220329

No effecting environmental conditions during analysis.

\*NMI is recognized by NIST through the Mutual Recognition Agreement (CIPM MRA).

CALGAZ calibration devices were found to meet all applicable requirements of the National Highway Traffic Safety Administration Model Specifications for calibrating units for breath alcohol testers.

Manufactured Date: December 14, 2023

APPROVED BY: 

"We certify that all the cylinders for the Lot numbers identified herein are manufactured and tested within the requirements of CFR 49 part 178.65 and that physical and chemical test reports are on file and copies will be furnished upon request."

**CALGAZ, a division of Airgas USA LLC**  
821 Chesapeake Drive, Cambridge, MD 21613-0149  
Phone: (410) 228-6400 Fax: (410) 228-4251

# CERTIFICATE OF ANALYSIS

## EBS - ETHANOL BREATH STANDARD

DEPT OF LAW AND PUBLIC SAFETY

Sales order: 120656632

Date: May 31, 2023

METHOD OF ANALYSIS: IR Breath Alcohol Analyzer  
ANALYTICAL ACCURACY: +/-0.002 BrAC or +/-2% whichever is greater.  
CALGAZ LOT#: 302-402757219  
ETHANOL IN NITROGEN

Product Expiration: May 30, 2026

COMPONENT	PPM	( BrAC )
ETHANOL	781.5PPM	(0.300)
NITROGEN	BAL	
AVERAGE ANALYTICAL VALUE	PPM	( BrAC )
ETHANOL	785.4	(0.301)

REFERENCE STANDARD	CYLINDER	CONCENTRATION PPM
N.M.I. TRACEABLE STANDARDS*	ND38424	260.7

\* CERTIFICATION TRACEABLE TO NATIONAL METROLOGY INSTITUTE TRACEABLE STANDARDS

### TRACEABILITY

#### Preparation:

Gas mixtures manufactured with balances calibrated by an ISO 17025 accredited company using NIST traceable weights and meets or exceeds the requirements of NIST Handbook 44.

Traceable certificate numbers 3445312 and 3398673.

#### Analytical:

Analytical Instruments Calibrated Using NMI Traceable Standards.

Certification Numbers: A679-20190918, D049803-20220329

No effecting environmental conditions during analysis.

\*NMI is recognized by NIST through the Mutual Recognition Agreement (CIPM MRA).

CALGAZ calibration devices were found to meet all applicable requirements of the National Highway Traffic Safety Administration Model Specifications for calibrating units for breath alcohol testers.

Manufactured Date: May 30, 2023

APPROVED BY: \_\_\_\_\_



"We certify that all the cylinders for the Lot numbers identified herein are manufactured and tested within the requirements of CFR 49 part 178.65 and that physical and chemical test reports are on file and copies will be furnished upon request."

**CALGAZ, a division of Airgas USA LLC**  
821 Chesapeake Drive, Cambridge, MD 21613-0149  
Phone: (410) 228-6400 Fax: (410) 228-4251

DEPARTMENT OF  
**Traffic and Public Safety**  
This is to certify that

**Anthony P. Rubbe**

**Breath Test Coordinator/Instructor**

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSIS PURSUANT TO CHAPTER 162 OF

TITLE 17 OF THE OPERATIONS OF THE **Alcotest 9510**

A METHOD TO CERTAIN INSTRUMENTS

GRIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 22nd DAY OF January

TWO THOUSAND AND Twenty Five

*[Signature]*  
COLONEL  
NEW JERSEY STATE POLICE

*[Signature]*  
ATTORNEY GENERAL  
STATE OF NEW JERSEY

DEPARTMENT OF  
**Traffic and Public Safety**  
This is to certify that

**Anthony P. Rubbe**

**New Jersey State Police**

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSIS PURSUANT TO CHAPTER 162 OF

TITLE 17 OF THE OPERATIONS OF THE **Alcotest 9510**

A METHOD TO DETERMINE INTOXICATION

GRIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 12th DAY OF October

TWO THOUSAND AND Twenty Three

*[Signature]*  
COLONEL  
NEW JERSEY STATE POLICE

*[Signature]*  
ATTORNEY GENERAL  
STATE OF NEW JERSEY

**ORIGINAL COURSE DATES**

	DATE	Refresher Course PLACE	INSTRUCTOR
1.			
2.			
3.			
4.			
5.			
6.			
7.			
8.			
9.			

S.P. 2938 (Rev. 01/18)

**ORIGINAL COURSE DATES**

	DATE	Refresher Course PLACE	INSTRUCTOR
1.	3-27-25	MCFA	<i>[Signature]</i>
2.			
3.			
4.			
5.			
6.			
7.			
8.			
9.			

S.P. 2938 (Rev. 10/22)

DEPARTMENT OF  
*Law and Public Safety*  
This is to certify that

**Nicholas E. Mimikos**

**New Jersey State Police**

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 142 OF

THE LAWS OF 1966 IN THE OPERATION OF THE Alcotest 9510

A METHOD TO DETERMINE INTOXICATION.  
 GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 8th DAY OF June

TWO THOUSAND AND Twenty One

  
 COLONEL  
 NEW JERSEY STATE POLICE

  
 ATTORNEY GENERAL  
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES

	DATE	Refresher Course PLACE	INSTRUCTOR
1.	7-14-23	Hamilton Tech	MF
2.			
3.			
4.			
5.			
6.			
7.			
8.			
9.			

S.P. 293B (Rev. 01/19)

DEPARTMENT OF  
*Law and Public Safety*  
This is to certify that

**Nicholas E. Mimikos**


**Breath Test Coordinator/Instructor**

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 142 OF

THE LAWS OF 1966 IN THE OPERATION OF THE Alcotest 9510

A METHOD TO DETERMINE INTOXICATION.  
 GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 8th DAY OF June

TWO THOUSAND AND Twenty One

  
 COLONEL  
 NEW JERSEY STATE POLICE

  
 ATTORNEY GENERAL  
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES

	DATE	Refresher Course PLACE	INSTRUCTOR
1.			
2.			
3.			
4.			
5.			
6.			
7.			
8.			
9.			

S.P. 293B (Rev. 01/18)

DEPARTMENT OF  
*Water and Public Safety*  
 This is to certify that

**Robert W. Waldrop**

**Breath Test Coordinator/Instructor**

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 142 OF

THE LAWS OF 1966 IN THE OPERATION OF THE Alcotest 9510

A METHOD TO DETERMINE INTOXICATION.

GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 25th DAY OF March

TWO THOUSAND AND Twenty Four

*[Signature]*  
 COLONEL  
 NEW JERSEY STATE POLICE

*[Signature]*  
 ATTORNEY GENERAL  
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES

DATE	Refresher Course PLACE	INSTRUCTOR
1.		
2.		
3.		
4.		
5.		
6.		
7.		
8.		
9.		

S.P. 2936 (Rev. 10/22)

DEPARTMENT OF  
*Water and Public Safety*  
 This is to certify that

**Robert W. Waldrop**

**New Jersey State Police**

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 142 OF

THE LAWS OF 1966 IN THE OPERATION OF THE Alcotest 9510

A METHOD TO DETERMINE INTOXICATION.

GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 28th DAY OF April

TWO THOUSAND AND Twenty Three

*[Signature]*  
 COLONEL  
 NEW JERSEY STATE POLICE

*[Signature]*  
 ATTORNEY GENERAL  
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES

DATE	Refresher Course PLACE	INSTRUCTOR
1.		
2.		
3.		
4.		
5.		
6.		
7.		
8.		
9.		

S.P. 2936 (Rev. 10/22)