

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00714	Calib. Date: 06/28/2023
Certification File No.:	00715	Calib. No.: 00053
Linearity File No.:	00716	Cert. Date: 06/28/2023
Solution File No.:	00726	Cert. No.: 00042
Sequential File No.:	00726	Lin. Date: 06/28/2023
		Lin. No.: 00042
		Soln. Date: 10/24/2023
		Soln. No.: 00288
		File Date: 10/24/2023

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUF S3-0061
Solution Control Lot:	23010	Expires: 01/11/2025
		Bottle No.: 0951

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	06:53D		
Control 1 EC	-.---%	06:54D	34.0°C	CTRL GAS SUPPLY
Control 1 IR	-.---%	06:54D	34.0°C	CTRL GAS SUPPLY
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GASSMAN

First Name: JASON

MI:

Signature: PFC  408

Badge No.: 408

Date: 10/24/2023

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00714	Calib. Date: 06/28/2023
Certification File No.:	00715	Calib. No.: 00053
Linearity File No.:	00716	Cert. Date: 06/28/2023
Solution File No.:	00727	Cert. No.: 00042
Sequential File No.:	00727	Lin. Date: 06/28/2023
		Lin. No.: 00042
		Soln. Date: 10/24/2023
		Soln. No.: 00289
		File Date: 10/24/2023

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUF S3-0061
Solution Control Lot:	23010	Expires: 01/11/2025
		Bottle No.: 1008

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:27D		
Control 1 EC	0.098%	09:28D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	09:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:29D		
Control 2 EC	0.099%	09:29D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	09:29D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:30D		
Control 3 EC	0.099%	09:31D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	09:31D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:31D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GASSMAN

First Name: JASON

MI:

Signature: PFC J Mh 408

Badge No.: 408

Date: 10/24/2023