

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0014
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 00714 Calib. Date: 06/28/2023 Calib. No.: 00053
Certification File No.: 00715 Cert. Date: 06/28/2023 Cert. No.: 00042
Linearity File No.: 00716 Lin. Date: 06/28/2023 Lin. No.: 00042
Solution File No.: 00729 Soln. Date: 11/15/2023 Soln. No.: 00290
Sequential File No.: 00729 File Date: 11/15/2023

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0061
Control Solution %: 0.100% Expires: 01/11/2025
Solution Control Lot: 23010 Bottle No.: 0943

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 08:54S | | |
| Control 1 EC | 0.099% | 08:55S | 33.9°C | *** TEST PASSED *** |
| Control 1 IR | 0.100% | 08:55S | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 08:56S | | |
| Control 2 EC | 0.099% | 08:56S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.100% | 08:56S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 08:57S | | |
| Control 3 EC | 0.100% | 08:58S | 33.9°C | *** TEST PASSED *** |
| Control 3 IR | 0.101% | 08:58S | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 08:58S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GASSMAN

First Name: JASON

MI:

Signature: _____

PFL / Mm 408

Badge No.: 408

Date: 11/15/2023