

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0014
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 00714 Calib. Date: 06/28/2023 Calib. No.: 00053
Certification File No.: 00715 Cert. Date: 06/28/2023 Cert. No.: 00042
Linearity File No.: 00716 Lin. Date: 06/28/2023 Lin. No.: 00042
Solution File No.: 00730 Soln. Date: 12/12/2023 Soln. No.: 00291
Sequential File No.: 00730 File Date: 12/12/2023

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0061
Control Solution %: 0.100% Expires: 04/12/2025
Solution Control Lot: 23110 Bottle No.: 0323

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:17S		
Control 1 EC	0.099%	12:18S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	12:18S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:18S		
Control 2 EC	0.099%	12:19S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	12:19S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:20S		
Control 3 EC	0.099%	12:20S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	12:20S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:21S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____

JFC. [Signature] #421

Badge No.: 421

Date: 12/12/2023