

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0014
Location:	SOUTH BRUNSWICK POLICE		
Calibration File No.:	00546	Calib. Date:	01/14/2019
Certification File No.:	00547	Cert. No.:	00040
Linearity File No.:	00548	Cert. Date:	01/14/2019
Solution File No.:	00556	Lin. No.:	00031
Sequential File No.:	00556	Soln. Date:	04/04/2019
		Soln. No.:	00220
		File Date:	04/04/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUH S3-0079
Solution Control Lot:	18150	Expires:	05/10/2020
		Bottle No.:	1060

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:14S		
Control 1 EC	0.102%	12:15S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	12:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:16S		
Control 2 EC	0.099%	12:16S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	12:16S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:17S		
Control 3 EC	0.100%	12:18S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	12:18S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:19S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: _____



Badge No.: 351

Date: 04/04/2019

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0014
Location:	SOUTH BRUNSWICK POLICE		
Calibration File No.:	00546	Calib. Date:	01/14/2019
Certification File No.:	00547	Cert. Date:	01/14/2019
Linearity File No.:	00548	Lin. Date:	01/14/2019
Solution File No.:	00555	Soln. Date:	04/04/2019
Sequential File No.:	00555	File Date:	04/04/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUH S3-0079
Solution Control Lot:	18150	Expires:	05/10/2020
		Bottle No.:	1178

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:09S		
Control 1 EC	0.000%	10:09S	34.0°C	CONTROL TEST FAILED
Control 1 IR	0.000%	10:09S	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	10:09S		
Control 2 EC	-.---%	10:10S	33.9°C	CTRL GAS SUPPLY
Control 2 IR	-.---%	10:10S	33.9°C	CTRL GAS SUPPLY
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: _____

SGT. Gary A. #351

Badge No.: 351

Date: 04/04/2019

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0014
Location:	SOUTH BRUNSWICK POLICE		
Calibration File No.:	00546	Calib. Date:	01/14/2019
Certification File No.:	00547	Cert. Date:	01/14/2019
Linearity File No.:	00548	Lin. Date:	01/14/2019
Solution File No.:	00554	Soln. Date:	04/04/2019
Sequential File No.:	00554	File Date:	04/04/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUH S3-0079
Solution Control Lot:	18150	Expires:	05/10/2020
		Bottle No.:	1178

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:11S		
Control 1 EC	-.---%	07:12S	21.8°C	SIMULATOR TEMP. ERROR
Control 1 IR	-.---%	07:12S	21.8°C	SIMULATOR TEMP. ERROR
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: _____

Sgt. Gary Holsten #351

Badge No.: 351

Date: 04/04/2019