

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00560	Calib. Date: 07/10/2019
Certification File No.:	00561	Calib. No.: 00041
Linearity File No.:	00562	Cert. Date: 07/10/2019
Solution File No.:	00563	Cert. No.: 00032
Sequential File No.:	00563	Lin. Date: 07/10/2019
		Lin. No.: 00032
		Soln. Date: 07/10/2019
		Soln. No.: 00224
		File Date: 07/10/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUH S3-0080
Solution Control Lot:	18150	Expires: 05/10/2020
		Bottle No.: 1016

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:15D		
Control 1 EC	0.100%	12:16D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	12:16D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:16D		
Control 2 EC	0.099%	12:17D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	12:17D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:18D		
Control 3 EC	0.099%	12:18D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	12:18D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:19D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDEE P2-022 #

Changed By:

Last Name: MIMIKOS

First Name: NICHOLAS

MI: E.

Signature: TPR. A. A #7413

Badge No.: 7413

Date: 07/10/2019