

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0014
Location:	SOUTH BRUNSWICK POLICE		
Calibration File No.:	00560	Calib. Date:	07/10/2019
Certification File No.:	00561	Cert. Date:	07/10/2019
Linearity File No.:	00562	Lin. Date:	07/10/2019
Solution File No.:	00567	Soln. Date:	10/31/2019
Sequential File No.:	00567	File Date:	10/31/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUH S3-0080
Solution Control Lot:	19020	Expires:	01/08/2021
		Bottle No.:	0270

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:02S		
Control 1 EC	0.099%	07:03S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	07:03S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:04S		
Control 2 EC	0.099%	07:04S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	07:04S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:05S		
Control 3 EC	0.099%	07:06S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	07:06S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:07S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____

Badge No.: 421

Date: 10/31/2019