

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0074
<b>Location:</b>	SOUTH BRUNSWICK POLICE		
<b>Calibration File No.:</b>	01250	<b>Calib. Date:</b>	06/28/2022
<b>Certification File No.:</b>	01251	<b>Cert. Date:</b>	06/28/2022
<b>Linearity File No.:</b>	01252	<b>Lin. Date:</b>	06/28/2022
<b>Solution File No.:</b>	01259	<b>Soln. Date:</b>	07/21/2022
<b>Sequential File No.:</b>	01259	<b>File Date:</b>	07/21/2022
<b>Calibrating Unit:</b>	WET	<b>Model No.:</b>	CU-34
<b>Control Solution %:</b>	0.100%	<b>Serial No.:</b>	DDUH S3-0079
<b>Solution Control Lot:</b>	21360	<b>Expires:</b>	09/01/2023
		<b>Bottle No.:</b>	1008

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:41D		
Control 1 EC	0.100%	07:42D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	07:42D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:43D		
Control 2 EC	0.099%	07:43D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	07:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:44D		
Control 3 EC	0.099%	07:45D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	07:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:45D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: GASSMAN

First Name: JASON

MI:

Signature: \_\_\_\_\_

*PFC J M 408*

Badge No.: 408

Date: 07/21/2022