

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0074
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 01282 Calib. Date: 12/12/2022 Calib. No.: 00051
Certification File No.: 01283 Cert. Date: 12/12/2022 Cert. No.: 00038
Linearity File No.: 01284 Lin. Date: 12/12/2022 Lin. No.: 00038
Solution File No.: 01287 Soln. Date: 01/05/2023 Soln. No.: 00278
Sequential File No.: 01287 File Date: 01/05/2023

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUH S3-0080
Control Solution %: 0.100% Expires: 09/01/2023
Solution Control Lot: 21360 Bottle No.: 1102

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:28S		
Control 1 EC	0.100%	08:29S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	08:29S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:30S		
Control 2 EC	0.100%	08:30S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:31S		
Control 3 EC	0.100%	08:31S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	08:31S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:32S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GASSMAN

First Name: JASON

MI:

Signature: _____

PFC JG 408

Badge No.: 408

Date: 01/05/2023