

# Calibrating Unit

## New Standard Solution Report

|                         |                        |              |              |
|-------------------------|------------------------|--------------|--------------|
| <b>Equipment</b>        | Alcotest 7110 MKIII-C  | Serial No.:  | ARNK-0074    |
| Location:               | SOUTH BRUNSWICK POLICE |              |              |
| Calibration File No.:   | 01282                  | Calib. Date: | 12/12/2022   |
| Certification File No.: | 01283                  | Cert. Date:  | 12/12/2022   |
| Linearity File No.:     | 01284                  | Lin. Date:   | 12/12/2022   |
| Solution File No.:      | 01289                  | Soln. Date:  | 02/02/2023   |
| Sequential File No.:    | 01289                  | File Date:   | 02/02/2023   |
| Calibrating Unit:       | WET                    | Model No.:   | CU-34        |
| Control Solution %:     | 0.100%                 | Serial No.:  | DDUH S3-0080 |
| Solution Control Lot:   | 22300                  | Expires:     | 08/09/2024   |
|                         |                        | Bottle No.:  | 1189         |

| Function          | Result<br>%BAC | Time<br>HH:MM | Temperature<br>Simulator (°C) | Comment(s)<br>or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000%         | 14:42S        |                               |                           |
| Control 1 EC      | 0.099%         | 14:43S        | 33.9°C                        | *** TEST PASSED ***       |
| Control 1 IR      | 0.100%         | 14:43S        | 33.9°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 14:44S        |                               |                           |
| Control 2 EC      | 0.099%         | 14:44S        | 33.9°C                        | *** TEST PASSED ***       |
| Control 2 IR      | 0.100%         | 14:44S        | 33.9°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 14:45S        |                               |                           |
| Control 3 EC      | 0.100%         | 14:46S        | 33.9°C                        | *** TEST PASSED ***       |
| Control 3 IR      | 0.100%         | 14:46S        | 33.9°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 14:46S        |                               |                           |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: GASSMAN

First Name: JASON

MI:

Signature: \_\_\_\_\_

*PFC J M 408*

Badge No.: 408

Date: 02/02/2023