

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0014
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 00684 Calib. Date: 12/12/2022 Calib. No.: 00051
Certification File No.: 00685 Cert. Date: 12/12/2022 Cert. No.: 00040
Linearity File No.: 00686 Lin. Date: 12/12/2022 Lin. No.: 00040
Solution File No.: 00694 Soln. Date: 02/06/2023 Soln. No.: 00278
Sequential File No.: 00694 File Date: 02/06/2023

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0060
Control Solution %: 0.100% Expires: 08/09/2024
Solution Control Lot: 22300 Bottle No.: 0202

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:34S		
Control 1 EC	0.100%	07:35S	33.8°C	*** TEST PASSED ***
Control 1 IR	0.099%	07:35S	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:35S		
Control 2 EC	0.099%	07:36S	33.8°C	*** TEST PASSED ***
Control 2 IR	0.099%	07:36S	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:37S		
Control 3 EC	0.099%	07:37S	33.8°C	*** TEST PASSED ***
Control 3 IR	0.099%	07:37S	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:38S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GASSMAN

First Name: JASON

MI:

Signature: _____

PFC J M 408

Badge No.: 408

Date: 02/06/2023