

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00684	Calib. No.: 00051
Certification File No.:	00685	Cert. No.: 00040
Linearity File No.:	00686	Lin. No.: 00040
Solution File No.:	00688	Soln. No.: 00274
Sequential File No.:	00688	File Date: 01/09/2023
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUF S3-0060
Solution Control Lot:	21360	Expires: 09/01/2023
		Bottle No.: 1105

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:07S		
Control 1 EC	-.---%	08:07S	33.8°C	CTRL GAS SUPPLY
Control 1 IR	-.---%	08:07S	33.8°C	CTRL GAS SUPPLY
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GASSMAN

First Name: JASON

MI:

Signature:  _____

Badge No.: 408

Date: 01/09/2023

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0014
Location:	SOUTH BRUNSWICK POLICE		
Calibration File No.:	00684	Calib. Date:	12/12/2022
Certification File No.:	00685	Cert. Date:	12/12/2022
Linearity File No.:	00686	Lin. Date:	12/12/2022
Solution File No.:	00689	Soln. Date:	01/09/2023
Sequential File No.:	00689	File Date:	01/09/2023

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUF S3-0060
Control Solution %:	0.100%			Expires:	08/09/2024
Solution Control Lot:	22300			Bottle No.:	0310

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:45S		
Control 1 EC	-.---%	11:45S	24.8°C	SIMULATOR TEMP. ERROR
Control 1 IR	-.---%	11:45S	24.8°C	SIMULATOR TEMP. ERROR
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GASSMAN

First Name: JASON

MI:

Signature: _____

Badge No.: 408

Date: 01/09/2023

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0014
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 00684 Calib. Date: 12/12/2022 Calib. No.: 00051
Certification File No.: 00685 Cert. Date: 12/12/2022 Cert. No.: 00040
Linearity File No.: 00686 Lin. Date: 12/12/2022 Lin. No.: 00040
Solution File No.: 00690 Soln. Date: 01/09/2023 Soln. No.: 00276
Sequential File No.: 00690 File Date: 01/09/2023

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0060
Control Solution %: 0.100% Expires: 08/09/2024
Solution Control Lot: 22300 Bottle No.: 0310

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:24S		
Control 1 EC	0.099%	14:25S	33.8°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:25S	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:25S		
Control 2 EC	0.100%	14:26S	33.8°C	*** TEST PASSED ***
Control 2 IR	0.099%	14:26S	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:27S		
Control 3 EC	0.100%	14:27S	33.8°C	*** TEST PASSED ***
Control 3 IR	0.100%	14:27S	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:28S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GASSMAN

First Name: JASON

MI:

Signature: _____



Badge No.: 408

Date: 01/09/2023

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00684	Calib. Date: 12/12/2022
Certification File No.:	00685	Calib. No.: 00051
Linearity File No.:	00686	Cert. Date: 12/12/2022
Solution File No.:	00691	Cert. No.: 00040
Sequential File No.:	00691	Lin. Date: 12/12/2022
		Lin. No.: 00040
		Soln. Date: 01/09/2023
		Soln. No.: 00277
		File Date: 01/09/2023
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUF S3-0060
Solution Control Lot:	22300	Expires: 08/09/2024
		Bottle No.: 1124

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:45S		
Control 1 EC	0.100%	15:46S	33.8°C	*** TEST PASSED ***
Control 1 IR	0.100%	15:46S	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:46S		
Control 2 EC	0.100%	15:47S	33.8°C	*** TEST PASSED ***
Control 2 IR	0.100%	15:47S	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:48S		
Control 3 EC	0.101%	15:48S	33.8°C	*** TEST PASSED ***
Control 3 IR	0.100%	15:48S	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:49S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GASSMAN

First Name: JASON

MI:

Signature: 

Badge No.: 408

Date: 01/09/2023