

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00684	Calib. Date: 12/12/2022
Certification File No.:	00685	Calib. No.: 00051
Linearity File No.:	00686	Cert. Date: 12/12/2022
Solution File No.:	00687	Cert. No.: 00040
Sequential File No.:	00687	Lin. Date: 12/12/2022
		Lin. No.: 00040
		Soln. Date: 12/12/2022
		Soln. No.: 00273
		File Date: 12/12/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUF S3-0060
Solution Control Lot:	22140	Expires: 04/27/2024
		Bottle No.: 0698

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:59S		
Control 1 EC	0.100%	15:00S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	15:00S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:01S		
Control 2 EC	0.100%	15:01S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	15:01S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:02S		
Control 3 EC	0.101%	15:03S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	15:03S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:03S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDEE P2 - 034 (8)

Changed By:

Last Name: GUARCELLO

First Name: COREY

MI: R.

Signature: 

Badge No.: 7035

Date: 12/12/2022