

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0074
Location:	SOUTH BRUNSWICK POLICE	Calib. No.:	00050
Calibration File No.:	01250	Cert. Date:	06/28/2022
Certification File No.:	01251	Lin. Date:	06/28/2022
Linearity File No.:	01252	Soln. Date:	11/08/2022
Solution File No.:	01276	File Date:	11/08/2022
Sequential File No.:	01276		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUH S3-0079
Control Solution %:	0.100%			Expires:	09/01/2023
Solution Control Lot:	21360			Bottle No.:	0901

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:48S		
Control 1 EC	0.099%	07:48S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	07:48S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:49S		
Control 2 EC	0.099%	07:50S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	07:50S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:50S		
Control 3 EC	0.099%	07:51S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	07:51S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:52S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: GASSMAN

First Name: JASON

MI:

Signature: \_\_\_\_\_

*PFC J M 408*

Badge No.: 408

Date: 11/08/2022