

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0074
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 01303 Calib. Date: 04/11/2023 Calib. No.: 00052
Certification File No.: 01304 Cert. Date: 04/11/2023 Cert. No.: 00039
Linearity File No.: 01305 Lin. Date: 04/11/2023 Lin. No.: 00039
Solution File No.: 01313 Soln. Date: 05/31/2023 Soln. No.: 00285
Sequential File No.: 01313 File Date: 05/31/2023

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUH S3-0079
Control Solution %: 0.100% Expires: 01/11/2025
Solution Control Lot: 23010 Bottle No.: 0862

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:21D		
Control 1 EC	0.101%	07:21D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	07:21D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:22D		
Control 2 EC	0.101%	07:23D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	07:23D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:24D		
Control 3 EC	0.101%	07:24D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	07:24D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:25D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GASSMAN

First Name: JASON

MI:

Signature: _____

PFC J M 408

Badge No.: 408

Date: 05/31/2023