

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARTL-0014  
**Location:** SOUTH BRUNSWICK POLICE  
**Calibration File No.:** 00701 **Calib. Date:** 04/11/2023 **Calib. No.:** 00052  
**Certification File No.:** 00702 **Cert. Date:** 04/11/2023 **Cert. No.:** 00041  
**Linearity File No.:** 00703 **Lin. Date:** 04/11/2023 **Lin. No.:** 00041  
**Solution File No.:** 00710 **Soln. Date:** 06/07/2023 **Soln. No.:** 00283  
**Sequential File No.:** 00710 **File Date:** 06/07/2023

**Calibrating Unit:** WET **Model No.:** CU-34 **Serial No.:** DDUF S3-0061  
**Control Solution %:** 0.100% **Expires:** 01/11/2025  
**Solution Control Lot:** 23010 **Bottle No.:** 1021

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:03D		
Control 1 EC	0.101%	08:04D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	08:04D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:04D		
Control 2 EC	0.101%	08:05D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	08:05D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:06D		
Control 3 EC	0.102%	08:06D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	08:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:07D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: GASSMAN

First Name: JASON

MI:

Signature: PFC J M 408

Badge No.: 408

Date: 06/07/2023